



SEA-BIRD
SCIENTIFIC

SBE Sea-Bird
Electronics

Sea-Bird Electronics
13431 NE 20th Street
Bellevue, Washington
98005 USA

Tel: +1 425-643-9866
seabird@seabird.com
www.seabird.com

SBE41-CP ALACE

Instrument Configuration

Instrument Serial Number: 41-8276
Instrument Firmware Version: V 7.2.5
Zero Conductivity Frequency: 2560.28
Communications Format: RS232
Communications Settings: 9600 baud, 8 Data Bits, No Parity

Installed Devices/Sensors

<i>Data Format</i>	<i>Measurement</i>	<i>Sensor Type</i>	<i>Serial Number</i>	<i>Rating</i>
Count	Temperature	Internal	N/A	N/A
Frequency	Conductivity	Internal	N/A	N/A
Count	Pressure	Kistler	4924213	2000m(2000 dBar)

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 8276
CALIBRATION DATE: 18-Mar-16

SBE 41 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:


a0 = -8.427755e-004
a1 = 2.877910e-004
a2 = -3.503033e-006
a3 = 1.425542e-007

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	17730026.4	1.0001	0.0001
4.5000	15125666.9	4.4999	-0.0001
14.9999	9578257.2	15.0000	0.0001
18.5000	8276833.9	18.5000	0.0000
23.9940	6620802.5	23.9940	-0.0000
29.0000	5435383.6	28.9999	-0.0001
32.5000	4750885.7	32.5001	0.0001

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature



Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 8276
CALIBRATION DATE: 18-Mar-16

SBE 41 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.921046e-001
h = 1.520783e-001
i = -4.231860e-004
j = 5.404406e-005

CPcor = -9.5700e-008
CTcor = 3.2500e-006
WBOTC = 3.9054e-007


BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2560.28	0.00000	0.00000
1.0000	34.7471	2.97061	5117.26	2.97062	0.00001
4.5000	34.7271	3.27714	5311.02	3.27713	-0.00001
14.9999	34.6841	4.25710	5887.15	4.25710	0.00000
18.5000	34.6750	4.60164	6076.47	4.60165	0.00000
23.9940	34.6650	5.15798	6369.98	5.15798	-0.00000
29.0000	34.6595	5.67952	6633.01	5.67951	-0.00001
32.5000	34.6565	6.05127	6814.10	6.05127	0.00000

$f = \text{Instrument Output(Hz)} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$

t = temperature (°C); p = pressure (decibars); $\delta = \text{CTcor}$; $\epsilon = \text{CPcor}$;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity



Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 8276

CALIBRATION DATE: 14-Mar-16

SBE 41 PRESSURE CALIBRATION DATA

2900 psia S/N 4924213

COEFFICIENTS:

PA0 =	-3.606343e-001	PTCA0 =	-1.259051e+004
PA1 =	3.905498e-004	PTCA1 =	-2.176089e+002
PA2 =	8.636331e-014	PTCA2 =	8.099082e+000
PTHA0 =	3.197753e+002	PTCB0 =	1.033414e+002
PTHA1 =	-9.066957e-005	PTCB1 =	-1.027111e-002
PTHA2 =	2.818421e-012	PTCB2 =	0.000000e+000

PRESSURE SPAN CALIBRATION

THERMAL CORRECTION

PRESSURE (PSIA)	INSTRUMENT OUTPUT (counts)	THERMISTOR OUTPUT (counts)	COMPUTED PRESSURE (PSIA)	RESIDUAL (%FSR)	TEMP (°C)	THERMISTOR OUTPUT (counts)	INSTRUMENT OUTPUT (counts)
14.55	24696.4	3718323.0	14.59	0.00	32.50	3563082.40	29785.90
591.43	1498309.9	3716119.4	591.55	0.00	29.00	3612589.40	28834.90
1168.28	2970564.2	3715098.4	1168.36	0.00	23.99	3683982.00	27764.28
1745.15	4442008.0	3714257.4	1745.24	0.00	18.50	3763065.00	27049.81
2321.96	5912460.6	3713571.2	2322.11	0.01	15.00	3813349.20	26878.59
2898.78	7381211.0	3712974.6	2898.69	-0.00	4.50	3966289.00	27496.88
2321.93	5911907.9	3712998.4	2321.90	-0.00	1.00	4017410.60	28112.09
1745.19	4441715.8	3712892.4	1745.14	-0.00			
1168.17	2969712.1	3712821.6	1168.04	-0.00			
591.42	1497335.0	3712627.2	591.17	-0.01			
14.55	24743.1	3712107.0	14.59	0.00			

TEMPERATURE (°C)	SPAN (mV)
-4.99	103.39
35.48	102.98

y = thermistor output (counts)

$$t = PTHA0 + PTHA1 * y + PTHA2 * y^2$$

$$x = \text{instrument output} - PTCA0 - PTCA1 * t - PTCA2 * t^2$$

$$n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * t^2)$$

$$\text{pressure (PSIA)} = PA0 + PA1 * n + PA2 * n^2$$

$$\text{Residual (\%FSR)} = (\text{computed pressure} - \text{true pressure}) * 100 / \text{Full Scale Range}$$

Date, Offset (%FSR)

● 14-Mar-16 -0.00

