

SBE 41 CERTIFICATES

Instrument Configuration.....	1
DC - Coefficient Output File.....	2
Temperature Calibration Sheet.....	3
Conductivity Calibration Sheet.....	4
Pressure Calibration Sheet.....	5

SEA-BIRD ELECTRONICS, INC.
13431 NE 20th Street
Bellevue, Washington 98005 USA
Phone: (425) 643 9866
Fax: (425) 643 9954
Email: seabird@seabird.com

SBE 41 Instrument Configuration

Model Number: SBE 41

Serial Number: 41-7172

Part Number: 41CP.FA200

Description: METOCEAN Standard Configuration

Firmware Version: 3.0C

Pressure Type: Kistler

Pressure Range: 2000 dBar

Pressure Serial Number: 4669401

SBE 41 ALACE-CP V 3.0C SERIAL NO. 7172
TEMPERATURE: 15-APR-15
TA0 = 5.096223E-05
TA1 = 2.690416E-04
TA2 = -2.034793E-06
TA3 = 1.411487E-07
CONDUCTIVITY: 15-APR-15
G = -9.785383E-01
H = 1.327226E-01
I = -3.295651E-04
J = 4.120937E-05
CPCOR = -9.570000E-08
CTCOR = 3.250000E-06
WBOTC = 2.755721E-07
PRESSURE S/N = 4669401, RANGE = 2900 PSIA: 06-APR-15
PA0 = -1.511758E-02
PA1 = 1.394180E-01
PA2 = 1.589426E-08
PTCA0 = -1.213198E+02
PTCA1 = -5.259296E-01
PTCA2 = 2.273958E-02
PTCB0 = 1.034770E+02
PTCB1 = -1.054279E-02
PTCB2 = 0.000000E+00
PTHA0 = -9.637022E+01
PTHA1 = 4.067486E-02
PTHA2 = 9.971424E-07
POFFSET = 0.000000E+00

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 7172
CALIBRATION DATE: 15-Apr-15

SBE 41 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:


a0 = 5.096223e-005
a1 = 2.690416e-004
a2 = -2.034793e-006
a3 = 1.411487e-007

BATH TEMP (ITS-90)	INSTRUMENT OUTPUT	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
1.0000	700384.9	1.0001	0.0001
4.5000	597827.4	4.4999	-0.0001
15.0000	379153.2	15.0000	-0.0000
18.5000	327794.1	18.5002	0.0002
23.9940	262404.7	23.9940	0.0000
29.0000	215566.1	28.9997	-0.0003
32.5000	188499.6	32.5002	0.0002

Temperature ITS-90 = $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$ (°C)

Residual = instrument temperature - bath temperature

n = instrument output



Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 7172
CALIBRATION DATE: 15-Apr-15

SBE 41 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.785383e-001
h = 1.327226e-001
i = -3.295651e-004
j = 4.120937e-005

CPcor = -9.5700e-008
CTcor = 3.2500e-006
WBOTC = 2.7557e-007


BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (Hz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2721.36	0.00000	0.00000
1.0000	34.7290	2.96921	5465.61	2.96922	0.00001
4.5000	34.7091	3.27561	5673.18	3.27560	-0.00001
15.0000	34.6665	4.25518	6290.21	4.25516	-0.00001
18.5000	34.6572	4.59954	6492.90	4.59953	-0.00000
23.9940	34.6471	5.15561	6807.15	5.15562	0.00001
29.0000	34.6416	5.67692	7088.71	5.67693	0.00001
32.5000	34.6390	6.04856	7282.56	6.04855	-0.00001

$$f = \text{INST FREQ} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$$

$$\text{Conductivity} = (g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p) \text{ Siemens / meter}$$

t = temperatur e[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = instrument conductivity - bath conductivity



Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 7172
CALIBRATION DATE: 06-Apr-15

SBE 41 PRESSURE CALIBRATION DATA
2900 psia S/N 4669401

COEFFICIENTS:

PA0 = -1.511758e-002	PTCA0 = -1.213198e+002
PA1 = 1.394180e-001	PTCA1 = -5.259296e-001
PA2 = 1.589426e-008	PTCA2 = 2.273958e-002
PTHA0 = -9.637022e+001	PTCB0 = 1.034770e+002
PTHA1 = 4.067486e-002	PTCB1 = -1.054279e-002
PTHA2 = 9.971424e-007	PTCB2 = 0.000000e+000

PRESSURE SPAN CALIBRATION					THERMAL CORRECTION		
PRESSURE PSIA	INST OUTPUT	THERMISTOR OUTPUT	COMPUTED PRESSURE	ERROR %FS	TEMP ITS90	PRESS TEMP	INST OUTPUT
14.53	-17.5	2747.9	14.51	-0.00	32.50	2954.20	-4.22
592.33	4117.6	2751.9	592.65	0.01	29.00	2879.10	-7.27
1169.94	8243.3	2753.3	1170.04	0.00	23.99	2771.20	-10.80
1747.62	12366.4	2755.8	1747.62	-0.00	18.50	2651.70	-13.10
2325.25	16486.3	2757.5	2325.30	0.00	15.00	2575.30	-13.89
2902.92	20602.4	2758.9	2903.00	0.00	4.50	2345.00	-13.05
2325.26	16485.4	2759.4	2325.19	-0.00	1.00	2267.90	-11.70
1747.66	12365.4	2760.1	1747.50	-0.01			
1170.04	8242.0	2761.0	1169.87	-0.01	TEMP(ITS90)	SPAN(mV)	
592.08	4114.0	2761.9	592.14	0.00	-5.38	103.53	
14.53	-18.1	2763.6	14.37	-0.01	35.52	103.10	

$$y = \text{thermistor output}; t = \text{PTHA0} + \text{PTHA1} * y + \text{PTHA2} * y^2$$

$$x = \text{pressure output} - \text{PTCA0} - \text{PTCA1} * t - \text{PTCA2} * t^2$$

$$n = x * \text{PTCB0} / (\text{PTCB0} + \text{PTCB1} * t + \text{PTCB2} * t^2)$$

$$\text{pressure (psia)} = \text{PA0} + \text{PA1} * n + \text{PA2} * n^2$$

