

SBE 41 CERTIFICATES

Instrument Configuration.....	1
DC - Coefficient Output File.....	2
Temperature Calibration Sheet.....	3
Conductivity Calibration Sheet.....	4
Pressure Calibration Sheet.....	5

SEA-BIRD ELECTRONICS, INC.
13431 NE 20th Street
Bellevue, Washington 98005 USA
Phone: (425) 643 9866
Fax: (425) 643 9954
Email: seabird@seabird.com

SBE 41 Instrument Configuration

Model Number: SBE 41

Serial Number: 41-7259

Part Number: 41CP.FA200

Description: METOCEAN Standard Configuration

Firmware Version: 3.0C

Pressure Type: Kistler

Pressure Range: 2000 dBar

Pressure Serial Number: 4645067

SBE 41 ALACE-CP V 3.0C SERIAL NO. 7259

TEMPERATURE: 20-MAY-15

TA0 = 1.020242E-04

TA1 = 2.587381E-04

TA2 = -1.278705E-06

TA3 = 1.199420E-07

CONDUCTIVITY: 20-MAY-15

G = -9.805769E-01

H = 1.341203E-01

I = -2.940315E-04

J = 3.827573E-05

CPCOR = -9.570001E-08

CTCOR = 3.250000E-06

WBOTC = 3.998434E-07

PRESSURE S/N = 4645067, RANGE = 2900 PSIA: 12-MAY-15

PA0 = -8.174979E-01

PA1 = 1.375609E-01

PA2 = 9.240087E-09

PTCA0 = 1.663845E+02

PTCA1 = -7.508758E-01

PTCA2 = 2.160668E-02

PTCB0 = 1.064790E+02

PTCB1 = -1.115113E-02

PTCB2 = 0.000000E+00

PTHA0 = -9.787642E+01

PTHA1 = 4.087029E-02

PTHA2 = 1.046597E-06

POFFSET = 0.000000E+00

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 7259
CALIBRATION DATE: 20-May-15

SBE 41 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:


a0 = 1.020242e-004
a1 = 2.587381e-004
a2 = -1.278705e-006
a3 = 1.199420e-007

BATH TEMP (ITS-90)	INSTRUMENT OUTPUT	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
1.0000	706105.4	1.0000	-0.0000
4.4999	602242.9	4.4999	0.0000
15.0000	381126.9	15.0000	-0.0000
18.5000	329277.5	18.5001	0.0001
23.9940	263318.6	23.9939	-0.0001
29.0000	216114.0	29.0001	0.0001
32.5000	188865.5	32.5000	-0.0000

Temperature ITS-90 = $1 / \{ a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)] \} - 273.15$ (°C)

Residual = instrument temperature - bath temperature

n = instrument output



Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 7259
CALIBRATION DATE: 20-May-15

SBE 41 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.805769e-001
h = 1.341203e-001
i = -2.940315e-004
j = 3.827573e-005

CPcor = -9.5700e-008
CTcor = 3.2500e-006
WBOTC = 3.9984e-007


BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (Hz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2709.12	0.00000	0.00000
1.0000	34.9777	2.98844	5449.42	2.98844	-0.00000
4.4999	34.9584	3.29680	5656.59	3.29680	0.00001
15.0000	34.9158	4.28252	6272.33	4.28252	-0.00001
18.5000	34.9062	4.62900	6474.59	4.62900	-0.00000
23.9940	34.8955	5.18848	6788.14	5.18849	0.00001
29.0000	34.8890	5.71288	7069.06	5.71289	0.00001
32.5000	34.8835	6.08638	7262.31	6.08637	-0.00001

$$f = \text{INST FREQ} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$$

$$\text{Conductivity} = (g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p) \text{ Siemens / meter}$$

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = instrument conductivity - bath conductivity



Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 7259
CALIBRATION DATE: 12-May-15

SBE 41 PRESSURE CALIBRATION DATA
2900 psia S/N 4645067

COEFFICIENTS:

PA0 = -8.174979e-001	PTCA0 = 1.663845e+002
PA1 = 1.375609e-001	PTCA1 = -7.508758e-001
PA2 = 9.240087e-009	PTCA2 = 2.160668e-002
PTHA0 = -9.787642e+001	PTCB0 = 1.064790e+002
PTHA1 = 4.087029e-002	PTCB1 = -1.115113e-002
PTHA2 = 1.046597e-006	PTCB2 = 0.000000e+000

PRESSURE SPAN CALIBRATION					THERMAL CORRECTION		
PRESSURE PSIA	INST OUTPUT	THERMISTOR OUTPUT	COMPUTED PRESSURE	ERROR %FS	TEMP ITS90	PRESS TEMP	INST OUTPUT
14.52	272.3	2753.7	14.61	0.00	32.50	2964.90	281.20
592.45	4461.9	2757.0	592.47	0.00	29.00	2890.40	279.18
1169.90	8646.5	2758.1	1169.99	0.00	23.99	2783.50	277.07
1747.36	12828.9	2759.2	1747.53	0.01	18.50	2665.50	276.24
2324.84	17007.9	2760.9	2324.94	0.00	15.00	2590.10	276.41
2902.23	21183.1	2761.8	2902.14	-0.00	4.50	2361.90	279.90
2324.82	17007.0	2761.7	2324.82	0.00	1.00	2285.60	281.95
1747.40	12827.4	2762.0	1747.34	-0.00			
1169.83	8644.3	2762.3	1169.70	-0.00	TEMP(ITS90)	SPAN(mV)	
592.35	4459.3	2762.2	592.12	-0.01	-3.92	106.52	
14.52	272.1	2764.1	14.56	0.00	35.78	106.08	

$$y = \text{thermistor output}; t = \text{PTHA0} + \text{PTHA1} * y + \text{PTHA2} * y^2$$

$$x = \text{pressure output} - \text{PTCA0} - \text{PTCA1} * t - \text{PTCA2} * t^2$$

$$n = x * \text{PTCB0} / (\text{PTCB0} + \text{PTCB1} * t + \text{PTCB2} * t^2)$$

$$\text{pressure (psia)} = \text{PA0} + \text{PA1} * n + \text{PA2} * n^2$$

