



**SEA-BIRD**  
SCIENTIFIC

## SBE41-CP ALACE

### Instrument Configuration

Instrument Serial Number: 41-10061  
Instrument Firmware Version: V 7.2.5  
Zero Conductivity Frequency: 2642.63  
Communications Format: RS232  
Communications Settings: 9600 baud, 8 Data Bits, No Parity

### Installed Devices/Sensors

<i>Data Format</i>	<i>Measurement</i>	<i>Sensor Type</i>	<i>Serial Number</i>	<i>Rating</i>
Count	Temperature	Internal	N/A	N/A
Frequency	Conductivity	Internal	N/A	N/A
Count	Pressure	Druck	10717932	2000m(2000 dBar)



Sea-Bird Scientific  
 13431 NE 20<sup>th</sup> Street  
 Bellevue, WA 98005  
 USA

+1 425-643-9866  
 seabird@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 10061  
 CALIBRATION DATE: 29-Aug-17

SBE 41 TEMPERATURE CALIBRATION DATA  
 ITS-90 TEMPERATURE SCALE

COEFFICIENTS:


a0 = -8.388020e-004  
 a1 = 2.953157e-004  
 a2 = -3.942334e-006  
 a3 = 1.523585e-007

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0001	14783644.6	1.0001	-0.0000
4.5000	12610182.4	4.5000	0.0000
15.0000	7982264.0	15.0000	0.0000
18.5002	6896917.0	18.5001	-0.0001
23.9940	5516117.3	23.9941	0.0001
29.0000	4527941.2	29.0000	0.0000
32.5000	3957457.4	32.5000	-0.0000

n = Instrument Output (counts)

$$\text{Temperature ITS-90 (°C)} = 1 / \{ a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)] \} - 273.15$$

Residual (°C) = instrument temperature - bath temperature





Sea-Bird Scientific  
 13431 NE 20<sup>th</sup> Street  
 Bellevue, WA 98005  
 USA

+1 425-643-9866  
 seabird@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 10061  
 CALIBRATION DATE: 29-Aug-17

SBE 41 CONDUCTIVITY CALIBRATION DATA  
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.007762e+000      CPcor = -9.5700e-008  
 h = 1.449836e-001      CTcor = 3.2500e-006  
 i = -3.940413e-004      WBOTC = 5.8966e-006  
 j = 4.950174e-005


BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2642.63	0.00000	0.00000
1.0001	34.6633	2.96414	5246.83	2.96414	0.00000
4.5000	34.6442	3.27008	5444.62	3.27008	-0.00000
15.0000	34.6021	4.24811	6032.76	4.24811	0.00000
18.5002	34.5933	4.59199	6226.06	4.59197	-0.00001
23.9940	34.5838	5.14723	6525.79	5.14725	0.00001
29.0000	34.5786	5.66775	6794.40	5.66775	0.00000
32.5000	34.5753	6.03870	6979.30	6.03870	-0.00000

$$f = \text{Instrument Output(Hz)} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$$

t = temperature (°C); p = pressure (decibars);  $\delta$  = CTcor;  $\epsilon$  = CPcor;

$$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$$

Residual (Siemens/meter) = instrument conductivity - bath conductivity





Sea-Bird Scientific  
 13431 NE 20<sup>th</sup> Street  
 Bellevue, WA 98005  
 USA

+1 425-643-9866  
 seabird@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 10061  
 CALIBRATION DATE: 22-Aug-17

SBE 41 PRESSURE CALIBRATION DATA  
 2900 psia S/N 10717932

COEFFICIENTS:

PA0 =	6.818914e-001	PTCA0 =	-1.683027e+003
PA1 =	3.932316e-004	PTCA1 =	1.140841e+002
PA2 =	-2.903172e-013	PTCA2 =	-1.538951e+000
PTHA0 =	3.343360e+002	PTCB0 =	2.516387e+001
PTHA1 =	-6.346525e-005	PTCB1 =	-2.487562e-005
PTHA2 =	-1.370566e-012	PTCB2 =	0.000000e+000

PRESSURE SPAN CALIBRATION

THERMAL CORRECTION

PRESSURE (PSIA)	INSTRUMENT OUTPUT (counts)	THERMISTOR OUTPUT (counts)	COMPUTED PRESSURE (PSIA)	RESIDUAL (%FSR)	TEMP (°C)	THERMISTOR OUTPUT (counts)	INSTRUMENT OUTPUT (counts)
14.61	35516.1	4495276.4	14.63	0.00	32.50	4347697.60	37144.10
591.45	1504334.9	4492632.8	591.57	0.00	29.00	4394065.60	37148.25
1168.64	2976905.3	4491507.8	1168.72	0.00	23.99	4460369.60	37032.97
1745.73	4452637.1	4490543.8	1745.85	0.00	18.50	4532922.20	36699.27
2322.82	5931357.3	4489344.2	2322.89	0.00	15.00	4578807.00	36407.71
2899.97	7413340.9	4488358.2	2899.92	-0.00	4.50	4716526.00	35570.97
2322.78	5931095.4	4488473.6	2322.78	0.00	1.00	4762566.80	35251.16
1745.78	4452311.8	4488385.8	1745.72	-0.00			
1168.66	2976106.0	4488329.4	1168.40	-0.01	TEMPERATURE (°C)	SPAN (mV)	
591.27	1503492.4	4488370.2	591.23	-0.00	-5.10	25.16	
14.61	35452.2	4487004.2	14.59	-0.00	35.10	25.16	

y = thermistor output (counts)

$$t = PTHA0 + PTHA1 * y + PTHA2 * y^2$$

$$x = \text{instrument output} - PTCA0 - PTCA1 * t - PTCA2 * t^2$$

$$n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * t^2)$$

$$\text{pressure (PSIA)} = PA0 + PA1 * n + PA2 * n^2$$

$$\text{Residual (\%FSR)} = (\text{computed pressure} - \text{true pressure}) * 100 / \text{Full Scale Range}$$

Date, Offset (%FSR)

● 22-Aug-17 0.00

