



**SEA-BIRD**  
SCIENTIFIC

## SBE41-CP ALACE

### Instrument Configuration

Instrument Serial Number: 41-10065  
Instrument Firmware Version: V 7.2.5  
Zero Conductivity Frequency: 2738.25  
Communications Format: RS232  
Communications Settings: 9600 baud, 8 Data Bits, No Parity

### Installed Devices/Sensors

<i>Data Format</i>	<i>Measurement</i>	<i>Sensor Type</i>	<i>Serial Number</i>	<i>Rating</i>
Count	Temperature	Internal	N/A	N/A
Frequency	Conductivity	Internal	N/A	N/A
Count	Pressure	Druck	10717051	2000m(2000 dBar)



Sea-Bird Scientific  
 13431 NE 20<sup>th</sup> Street  
 Bellevue, WA 98005  
 USA

+1 425-643-9866  
 seabird@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 10065  
 CALIBRATION DATE: 28-Aug-17

SBE 41 TEMPERATURE CALIBRATION DATA  
 ITS-90 TEMPERATURE SCALE

COEFFICIENTS:


a0 = -7.288436e-004  
 a1 = 2.742123e-004  
 a2 = -2.621181e-006  
 a3 = 1.255736e-007

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	14726801.1	1.0000	0.0000
4.5000	12565737.0	4.5000	-0.0000
15.0000	7961553.6	15.0001	0.0001
18.5000	6881072.3	18.5000	-0.0000
23.9940	5505834.8	23.9939	-0.0001
29.0000	4521103.5	29.0001	0.0001
32.5000	3952447.0	32.4999	-0.0001

n = Instrument Output (counts)

$$\text{Temperature ITS-90 (°C)} = 1 / \{ a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)] \} - 273.15$$

Residual (°C) = instrument temperature - bath temperature





Sea-Bird Scientific  
 13431 NE 20<sup>th</sup> Street  
 Bellevue, WA 98005  
 USA

+1 425-643-9866  
 seabird@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 10065  
 CALIBRATION DATE: 28-Aug-17

SBE 41 CONDUCTIVITY CALIBRATION DATA  
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.978756e-001                      CPcor = -9.5700e-008  
 h = 1.337074e-001                      CTcor = 3.2500e-006  
 i = -3.487843e-004                      WBOTC = 6.3297e-006  
 j = 4.192045e-005


BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2738.25	0.00000	0.00000
1.0000	34.8727	2.98033	5468.00	2.98031	-0.00001
4.5000	34.8529	3.28784	5674.86	3.28785	0.00001
15.0000	34.8105	4.27098	6289.84	4.27098	0.00000
18.5000	34.8014	4.61661	6491.90	4.61661	0.00000
23.9940	34.7917	5.17475	6805.20	5.17473	-0.00002
29.0000	34.7866	5.69800	7085.98	5.69802	0.00001
32.5000	34.7833	6.07089	7279.22	6.07088	-0.00000

$$f = \text{Instrument Output(Hz)} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$$

t = temperature (°C); p = pressure (decibars);  $\delta$  = CTcor;  $\epsilon$  = CPcor;

$$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$$

Residual (Siemens/meter) = instrument conductivity - bath conductivity





Sea-Bird Scientific  
 13431 NE 20<sup>th</sup> Street  
 Bellevue, WA 98005  
 USA

+1 425-643-9866  
 seabird@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 10065  
 CALIBRATION DATE: 22-Aug-17

SBE 41 PRESSURE CALIBRATION DATA  
 2900 psia S/N 10717051

COEFFICIENTS:

PA0 =	4.749831e-001	PTCA0 =	-3.773185e+003
PA1 =	3.932697e-004	PTCA1 =	4.405954e+001
PA2 =	-2.897809e-013	PTCA2 =	5.881639e-001
PTHA0 =	3.358224e+002	PTCB0 =	2.511350e+001
PTHA1 =	-6.262900e-005	PTCB1 =	-9.000000e-004
PTHA2 =	-1.525158e-012	PTCB2 =	0.000000e+000

PRESSURE SPAN CALIBRATION

THERMAL CORRECTION

PRESSURE (PSIA)	INSTRUMENT OUTPUT (counts)	THERMISTOR OUTPUT (counts)	COMPUTED PRESSURE (PSIA)	RESIDUAL (%FSR)	TEMP (°C)	THERMISTOR OUTPUT (counts)	INSTRUMENT OUTPUT (counts)
14.61	33456.9	4523357.6	14.65	0.00	32.50	4376663.60	35832.50
591.23	1500167.5	4519362.8	591.25	0.00	29.00	4422737.60	35613.30
1167.98	2970515.8	4517726.2	1168.03	0.00	23.99	4488410.40	35257.80
1744.83	4444208.7	4516485.6	1744.87	0.00	18.50	4560206.20	34813.10
2321.64	5921050.1	4515381.8	2321.69	0.00	15.00	4605978.00	34563.90
2898.46	7400931.2	4514340.8	2898.42	-0.00	4.50	4742545.00	34048.50
2321.63	5920910.5	4514289.2	2321.64	0.00	1.00	4787861.80	33851.20
1744.94	4444359.1	4514193.2	1744.94	-0.00			
1167.96	2970100.6	4514091.4	1167.87	-0.00			
591.10	1499587.6	4513753.8	591.02	-0.00	TEMPERATURE (°C)	SPAN (mV)	
14.61	33412.2	4512924.6	14.61	0.00	-5.00	25.12	
					35.00	25.08	

y = thermistor output (counts)

$$t = PTHA0 + PTHA1 * y + PTHA2 * y^2$$

$$x = \text{instrument output} - PTCA0 - PTCA1 * t - PTCA2 * t^2$$

$$n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * t^2)$$

$$\text{pressure (PSIA)} = PA0 + PA1 * n + PA2 * n^2$$

$$\text{Residual (\%FSR)} = (\text{computed pressure} - \text{true pressure}) * 100 / \text{Full Scale Range}$$

Date, Offset (%FSR)

● 22-Aug-17 0.00

