



**SEA-BIRD**  
SCIENTIFIC

## SBE41-CP ALACE

### Instrument Configuration

Instrument Serial Number: 41-10598  
Instrument Firmware Version: V 7.2.5  
Zero Conductivity Frequency: 2615.75  
Communications Format: RS232  
Communications Settings: 9600 baud, 8 Data Bits, No Parity

### Installed Devices/Sensors

<i>Data Format</i>	<i>Measurement</i>	<i>Sensor Type</i>	<i>Serial Number</i>	<i>Rating</i>
Count	Temperature	Internal	N/A	N/A
Frequency	Conductivity	Internal	N/A	N/A
Count	Pressure	Druck	10817714	2000m(2000 dBar)



Sea-Bird Scientific  
 13431 NE 20<sup>th</sup> Street  
 Bellevue, WA 98005  
 USA

+1 425-643-9866  
 seabird@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 10598  
 CALIBRATION DATE: 04-Mar-18

SBE 41 TEMPERATURE CALIBRATION DATA  
 ITS-90 TEMPERATURE SCALE

COEFFICIENTS:


a0 = -7.987911e-004  
 a1 = 2.822341e-004  
 a2 = -3.156637e-006  
 a3 = 1.352085e-007

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	16918099.2	1.0000	0.0000
4.5000	14429270.7	4.4999	-0.0001
15.0000	9130760.8	15.0000	0.0000
18.5000	7888394.0	18.5001	0.0001
23.9940	6307971.0	23.9940	-0.0000
29.0000	5177046.5	28.9999	-0.0001
32.5000	4524183.3	32.5001	0.0001

n = Instrument Output (counts)

$$\text{Temperature ITS-90 (°C)} = 1 / \{ a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)] \} - 273.15$$

Residual (°C) = instrument temperature - bath temperature





Sea-Bird Scientific  
 13431 NE 20<sup>th</sup> Street  
 Bellevue, WA 98005  
 USA

+1 425-643-9866  
 seabird@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 10598  
 CALIBRATION DATE: 04-Mar-18

SBE 41 CONDUCTIVITY CALIBRATION DATA  
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.015547e+000      CPcor = -9.5700e-008  
 h = 1.489495e-001      CTcor = 3.2500e-006  
 i = -3.182206e-004      WBOTC = -1.2742e-007  
 j = 4.504337e-005


BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2615.75	0.00000	0.00000
1.0000	34.7076	2.96756	5178.86	2.96756	-0.00000
4.5000	34.6885	3.27385	5373.77	3.27386	0.00000
15.0000	34.6472	4.25306	5953.49	4.25305	-0.00001
18.5000	34.6388	4.59736	6144.08	4.59735	-0.00000
23.9940	34.6296	5.15330	6439.64	5.15330	-0.00000
29.0000	34.6249	5.67449	6704.58	5.67450	0.00001
32.5000	34.6231	6.04610	6887.06	6.04609	-0.00001

$$f = \text{Instrument Output(Hz)} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$$

t = temperature (°C); p = pressure (decibars);  $\delta$  = CTcor;  $\epsilon$  = CPcor;

$$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$$

Residual (Siemens/meter) = instrument conductivity - bath conductivity





Sea-Bird Scientific  
 13431 NE 20<sup>th</sup> Street  
 Bellevue, WA 98005  
 USA

+1 425-643-9866  
 seabird@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 10598  
 CALIBRATION DATE: 08-Feb-18

SBE 41 PRESSURE CALIBRATION DATA  
 2900 psia S/N 10817714

COEFFICIENTS:

PA0 =	5.763526e-001	PTCA0 =	-1.413813e+003
PA1 =	3.951143e-004	PTCA1 =	8.732367e+001
PA2 =	-2.891035e-013	PTCA2 =	-9.025634e-001
PTHA0 =	3.505982e+002	PTCB0 =	3.094597e+005
PTHA1 =	-6.367266e-005	PTCB1 =	3.713674e+000
PTHA2 =	-1.645866e-012	PTCB2 =	-1.448924e-001

PRESSURE SPAN CALIBRATION

THERMAL CORRECTION

PRESSURE (PSIA)	INSTRUMENT OUTPUT (counts)	THERMISTOR OUTPUT (counts)	COMPUTED PRESSURE (PSIA)	RESIDUAL (%FSR)	TEMP (°C)	THERMISTOR OUTPUT (counts)	INSTRUMENT OUTPUT (counts)
14.68	35833.6	4617858.4	14.72	0.00	32.50	4477604.20	37593.00
591.94	1498592.1	4613772.4	592.00	0.00	29.00	4522169.40	37544.01
1169.26	2964697.5	4612525.6	1169.36	0.00	23.99	4585841.80	37365.72
1746.65	4434062.0	4611405.2	1746.76	0.00	18.50	4655435.20	37077.07
2324.05	5906644.0	4610338.0	2324.18	0.00	15.00	4699815.00	36802.70
2901.40	7381808.7	4609001.0	2901.35	-0.00	4.50	4831960.00	36079.80
2324.07	5906270.0	4608854.4	2324.03	-0.00	1.00	4876044.00	35873.89
1746.77	4433903.0	4608540.2	1746.70	-0.00			
1169.44	2964335.8	4608369.4	1169.21	-0.01			
591.85	1498192.2	4608297.0	591.83	-0.00			
14.68	35722.3	4607989.4	14.66	-0.00			

TEMPERATURE (°C)	SPAN
2.18	309467.16
23.04	309468.40
32.58	309426.96

y = thermistor output (counts)

$$t = PTHA0 + PTHA1 * y + PTHA2 * y^2$$

$$x = \text{instrument output} - PTCA0 - PTCA1 * t - PTCA2 * t^2$$

$$n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * t^2)$$

$$\text{pressure (PSIA)} = PA0 + PA1 * n + PA2 * n^2$$

$$\text{Residual (\%FSR)} = (\text{computed pressure} - \text{true pressure}) * 100 / \text{Full Scale Range}$$

Date, Offset (%FSR)

● 08-Feb-18 0.00

