



**SEA-BIRD**  
SCIENTIFIC

## SBE41-CP ALACE

### Instrument Configuration

Instrument Serial Number: 41-11572  
Instrument Firmware Version: V 7.2.5  
Zero Conductivity Frequency: 2571.73  
Communications Format: RS232  
Communications Settings: 9600 baud, 8 Data Bits, No Parity

### Installed Devices/Sensors

<i>Data Format</i>	<i>Measurement</i>	<i>Sensor Type</i>	<i>Serial Number</i>	<i>Rating</i>
Count	Temperature	Internal	N/A	N/A
Frequency	Conductivity	Internal	N/A	N/A
Count	Pressure	Druck	10997288	2000m(2000 dBar)



Sea-Bird Scientific  
 13431 NE 20<sup>th</sup> Street  
 Bellevue, WA 98005  
 USA

+1 425-643-9866  
 seabird@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 11572  
 CALIBRATION DATE: 01-Apr-19

SBE 41 TEMPERATURE CALIBRATION DATA  
 ITS-90 TEMPERATURE SCALE

COEFFICIENTS:


a0 = -8.764162e-004  
 a1 = 2.981127e-004  
 a2 = -4.012510e-006  
 a3 = 1.574562e-007

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	14174946.6	1.0000	-0.0000
4.5000	12119220.1	4.5000	-0.0000
15.0000	7722820.3	15.0000	0.0000
18.5000	6686900.1	18.5000	-0.0000
23.9940	5365355.7	23.9940	-0.0000
29.0000	4416624.1	29.0000	0.0000
32.5000	3867576.5	32.5000	-0.0000

n = Instrument Output (counts)

$$\text{Temperature ITS-90 (°C)} = 1 / \{ a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)] \} - 273.15$$

Residual (°C) = instrument temperature - bath temperature





Sea-Bird Scientific  
 13431 NE 20<sup>th</sup> Street  
 Bellevue, WA 98005  
 USA

+1 425-643-9866  
 seabird@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 11572  
 CALIBRATION DATE: 01-Apr-19

SBE 41 CONDUCTIVITY CALIBRATION DATA  
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.001414e+000      CPcor = -9.5700e-008  
 h = 1.518972e-001      CTcor = 3.2500e-006  
 i = -3.114094e-004      WBOTC = -4.5360e-007  
 j = 4.804153e-005


BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2571.73	0.00000	0.00000
1.0000	34.7502	2.97085	5119.47	2.97085	0.00000
4.5000	34.7309	3.27746	5312.75	3.27745	-0.00001
15.0000	34.6897	4.25772	5887.53	4.25773	0.00001
18.5000	34.6814	4.60240	6076.44	4.60241	0.00001
23.9940	34.6725	5.15898	6369.34	5.15895	-0.00003
29.0000	34.6684	5.68081	6631.90	5.68082	0.00001
32.5000	34.6662	6.05277	6812.66	6.05277	0.00000

$f = \text{Instrument Output(Hz)} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$

t = temperature (°C); p = pressure (decibars);  $\delta = \text{CTcor}$ ;  $\epsilon = \text{CPcor}$ ;

$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity





Sea-Bird Scientific  
 13431 NE 20<sup>th</sup> Street  
 Bellevue, WA 98005  
 USA

+1 425-643-9866  
 seabird@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 11572  
 CALIBRATION DATE: 18-Mar-19

SBE 41 PRESSURE CALIBRATION DATA  
 2900 psia S/N 10997288

COEFFICIENTS:

PA0 =	9.650893e-001	PTCA0 =	1.376927e+003
PA1 =	3.927441e-004	PTCA1 =	8.308822e+001
PA2 =	-2.833381e-013	PTCA2 =	1.393920e+000
PTHA0 =	3.512253e+002	PTCB0 =	3.132592e+005
PTHA1 =	-6.562070e-005	PTCB1 =	-1.007684e+001
PTHA2 =	-1.360759e-012	PTCB2 =	3.624837e-001

PRESSURE SPAN CALIBRATION

THERMAL CORRECTION

PRESSURE (PSIA)	INSTRUMENT OUTPUT (counts)	THERMISTOR OUTPUT (counts)	COMPUTED PRESSURE (PSIA)	RESIDUAL (%FSR)	TEMP (°C)	THERMISTOR OUTPUT (counts)	INSTRUMENT OUTPUT (counts)
14.70	38633.0	4585837.4	14.63	-0.00	32.50	4446990.00	41832.70
591.90	1510236.2	4582031.0	592.02	0.00	29.00	4492024.00	41379.42
1169.45	2985307.1	4580960.0	1169.55	0.00	23.99	4556248.60	40657.21
1746.86	4463149.4	4578742.8	1746.92	0.00	18.50	4626415.80	39772.27
2324.33	5944357.8	4577803.4	2324.36	0.00	15.00	4671378.40	39201.12
2901.80	7428821.0	4576944.2	2901.83	0.00	4.50	4805046.60	38093.79
2324.28	5943987.5	4577056.8	2324.21	-0.00	1.00	4849411.40	37893.04
1746.96	4463074.7	4577129.4	1746.88	-0.00			
1169.46	2984877.7	4577045.2	1169.36	-0.00			
591.79	1509555.9	4577072.8	591.73	-0.00			
14.70	38979.5	4576293.6	14.73	0.00			

TEMPERATURE (°C)	SPAN
0.94	313250.05
20.05	313202.93
32.54	313315.20

y = thermistor output (counts)

$$t = PTHA0 + PTHA1 * y + PTHA2 * y^2$$

$$x = \text{instrument output} - PTCA0 - PTCA1 * t - PTCA2 * t^2$$

$$n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * t^2)$$

$$\text{pressure (PSIA)} = PA0 + PA1 * n + PA2 * n^2$$

$$\text{Residual (\%FSR)} = (\text{computed pressure} - \text{true pressure}) * 100 / \text{Full Scale Range}$$

Date, Offset (%FSR)

● 18-Mar-19 -0.00

