



SEA-BIRD
SCIENTIFIC

SBE Sea-Bird
Electronics

Sea-Bird Electronics
13431 NE 20th Street
Bellevue, Washington
98005 USA

Tel: +1 425-643-9866
seabird@seabird.com
www.seabird.com

SBE 41-CP ALACE

Instrument Configuration

Instrument Serial Number: 41-7275
Instrument Firmware Version: ALACE-CP V 3.0C
Zero Conductivity Frequency: 2702.83

Installed Devices/Sensors

<i>Data Format</i>	<i>Measurement</i>	<i>Sensor Type</i>	<i>Serial Number</i>	<i>Rating</i>
Count	Temperature	Internal	N/A	N/A
Frequency	Conductivity	Internal	N/A	N/A
Count	Pressure	Kistler	4669454	2000m(2000 dBar)

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 7275
CALIBRATION DATE: 23-Jun-15

SBE 41 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:


a0 = 7.338958e-005
a1 = 2.713585e-004
a2 = -2.214963e-006
a3 = 1.469158e-007

BATH TEMP (ITS-90)	INSTRUMENT OUTPUT	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
1.0000	620975.9	1.0000	0.0000
4.5000	529959.3	4.5000	-0.0000
15.0000	335956.9	15.0001	0.0001
18.5000	290412.5	18.4999	-0.0001
23.9940	232428.1	23.9939	-0.0001
29.0000	190899.2	29.0002	0.0002
32.5000	166913.7	32.4999	-0.0001

Temperature ITS-90 = $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$ (°C)

Residual = instrument temperature - bath temperature

n = instrument output



Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 7275
CALIBRATION DATE: 23-Jun-15

SBE 41 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.878538e-001
h = 1.358743e-001
i = -3.580837e-004
j = 4.363115e-005

CPcor = -9.5700e-008
CTcor = 3.2500e-006
WBOTC = -6.1657e-008


BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (Hz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2702.82	0.00000	0.00000
1.0000	34.7969	2.97446	5413.33	2.97447	0.00001
4.5000	34.7773	3.28141	5618.58	3.28140	-0.00001
15.0000	34.7350	4.26269	6228.78	4.26268	-0.00001
18.5000	34.7260	4.60768	6429.27	4.60768	-0.00000
23.9940	34.7163	5.16477	6740.11	5.16478	0.00001
29.0000	34.7109	5.68700	7018.64	5.68701	0.00001
32.5000	34.7082	6.05927	7210.39	6.05926	-0.00001

$$f = \text{INST FREQ} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$$

$$\text{Conductivity} = (g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p) \text{ Siemens / meter}$$

t = temperatur e[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = instrument conductivity - bath conductivity



Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 7275
CALIBRATION DATE: 19-Jun-15

SBE 41 PRESSURE CALIBRATION DATA
2900 psia S/N 4669454

COEFFICIENTS:

PA0 = -1.091727e+000	PTCA0 = -1.908627e+002
PA1 = 1.399533e-001	PTCA1 = -8.185042e-001
PA2 = 1.124805e-008	PTCA2 = 2.134607e-002
PTHA0 = -9.627321e+001	PTCB0 = 1.030847e+002
PTHA1 = 4.047546e-002	PTCB1 = -8.465507e-003
PTHA2 = 1.027046e-006	PTCB2 = 0.000000e+000

PRESSURE SPAN CALIBRATION

PRESSURE PSIA	INST OUTPUT	THERMISTOR OUTPUT	COMPUTED PRESSURE	ERROR %FS
14.67	-85.7	2710.0	14.74	0.00
592.83	4036.3	2720.7	592.84	0.00
1170.50	8152.9	2722.8	1170.59	0.00
1748.26	12267.5	2724.4	1748.43	0.01
2326.01	16378.4	2726.1	2326.15	0.00
2903.64	20484.0	2728.2	2903.52	-0.00
2325.91	16376.5	2728.9	2325.91	-0.00
1748.23	12265.5	2729.7	1748.18	-0.00
1170.55	8151.4	2730.3	1170.40	-0.00
592.59	4032.5	2731.4	592.33	-0.01
14.67	-85.5	2733.8	14.76	0.00

THERMAL CORRECTION

TEMP ITS90	PRESS TEMP	INST OUTPUT
32.50	2959.10	-75.80
29.00	2884.20	-77.91
23.99	2775.90	-79.29
18.50	2656.70	-79.75
15.00	2580.00	-79.39
4.50	2349.50	-74.93
1.00	2272.40	-72.90

TEMP(ITS90)	SPAN(mV)
-4.38	103.12
37.48	102.77

$$y = \text{thermistor output}; t = \text{PTHA0} + \text{PTHA1} * y + \text{PTHA2} * y^2$$

$$x = \text{pressure output} - \text{PTCA0} - \text{PTCA1} * t - \text{PTCA2} * t^2$$

$$n = x * \text{PTCB0} / (\text{PTCB0} + \text{PTCB1} * t + \text{PTCB2} * t^2)$$

$$\text{pressure (psia)} = \text{PA0} + \text{PA1} * n + \text{PA2} * n^2$$

