



SEA-BIRD  
SCIENTIFIC

**SBE** Sea-Bird  
Electronics

Sea-Bird Electronics  
13431 NE 20<sup>th</sup> Street  
Bellevue, Washington  
98005 USA

Tel: +1 425-643-9866  
seabird@seabird.com  
www.seabird.com

## SBE41-CP ALACE

### Instrument Configuration

Instrument Serial Number: 41-7298  
Instrument Firmware Version: ALACE-CP V 3.0C  
Zero Conductivity Frequency: 2690.77  
Communications Format: RS232  
Communications Settings: 9600 baud, 8 Data Bits, No Parity

### Installed Devices/Sensors

<i>Data Format</i>	<i>Measurement</i>	<i>Sensor Type</i>	<i>Serial Number</i>	<i>Rating</i>
Count	Temperature	Internal	N/A	N/A
Frequency	Conductivity	Internal	N/A	N/A
Count	Pressure	Kistler	4645062	2000m(2000 dBar)

**CAUTION - This instrument is not intended for underwater use**

# Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 7298  
CALIBRATION DATE: 11-Jun-15

SBE 41 TEMPERATURE CALIBRATION DATA  
ITS-90 TEMPERATURE SCALE

## COEFFICIENTS:


a0 = 4.224866e-005  
a1 = 2.671954e-004  
a2 = -1.963730e-006  
a3 = 1.368336e-007

BATH TEMP (ITS-90)	INSTRUMENT OUTPUT	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
1.0000	780233.7	1.0000	0.0000
4.5000	665432.6	4.5000	-0.0000
15.0000	421042.3	15.0001	0.0001
18.5000	363743.0	18.5000	-0.0000
23.9940	290853.0	23.9939	-0.0001
29.0000	238692.6	29.0001	0.0001
32.5001	208586.4	32.5000	-0.0001

Temperature ITS-90 =  $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$  (°C)

Residual = instrument temperature - bath temperature

n = instrument output



# Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 7298  
 CALIBRATION DATE: 11-Jun-15

SBE 41 CONDUCTIVITY CALIBRATION DATA  
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

## COEFFICIENTS:

g = -9.856301e-001  
 h = 1.366503e-001  
 i = -2.983670e-004  
 j = 3.963815e-005

CPcor = -9.5700e-008  
 CTcor = 3.2500e-006  
 WBOTC = -5.8836e-007


BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (Hz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2690.76	0.00000	0.00000
1.0000	34.8502	2.97859	5395.11	2.97859	0.00000
4.5000	34.8302	3.28591	5599.77	3.28590	-0.00000
15.0000	34.7873	4.26843	6208.25	4.26844	0.00000
18.5000	34.7776	4.61379	6408.13	4.61378	-0.00001
23.9940	34.7670	5.17148	6718.06	5.17149	0.00001
29.0000	34.7610	5.69428	6995.77	5.69428	0.00000
32.5001	34.7573	6.06688	7186.95	6.06687	-0.00000

$$f = \text{INST FREQ} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$$

$$\text{Conductivity} = (g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p) \text{ Siemens / meter}$$

t = temperatur e[°C]; p = pressure[decibars];  $\delta$  = CTcor;  $\epsilon$  = CPcor;

Residual = instrument conductivity - bath conductivity



# Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 7298  
CALIBRATION DATE: 02-Jun-15

SBE 41 PRESSURE CALIBRATION DATA  
2900 psia S/N 4645062

## COEFFICIENTS:

PA0 =	9.977793e-002	PTCA0 =	-1.170266e+002
PA1 =	1.412982e-001	PTCA1 =	-5.626529e-001
PA2 =	7.951025e-009	PTCA2 =	2.539145e-002
PTHA0 =	-9.198606e+001	PTCB0 =	1.024663e+002
PTHA1 =	3.779574e-002	PTCB1 =	-8.607053e-003
PTHA2 =	1.161542e-006	PTCB2 =	0.000000e+000

## PRESSURE SPAN CALIBRATION

PRESSURE PSIA	INST OUTPUT	THERMISTOR OUTPUT	COMPUTED PRESSURE	ERROR %FS
14.61	-12.9	2809.1	14.74	0.00
592.79	4069.2	2812.5	592.80	0.00
1170.47	8147.5	2814.5	1170.61	0.00
1748.31	12224.7	2815.9	1748.52	0.01
2326.06	16298.4	2816.6	2326.22	0.01
2903.84	20368.8	2817.4	2903.71	-0.00
2326.03	16297.1	2817.9	2326.04	0.00
1748.18	12221.7	2818.3	1748.10	-0.00
1170.57	8145.8	2818.1	1170.37	-0.01
592.84	4067.5	2819.0	592.55	-0.01
14.61	-13.3	2819.9	14.64	0.00

## THERMAL CORRECTION

TEMP ITS90	PRESS TEMP	INST OUTPUT
32.50	3014.20	-0.87
29.00	2936.30	-4.06
23.99	2823.70	-7.79
18.50	2699.40	-11.14
15.00	2619.60	-12.03
4.50	2378.60	-11.27
1.00	2298.20	-9.69

TEMP(ITS90)	SPAN(mV)
-3.92	102.50
35.78	102.16

$$y = \text{thermistor output}; t = \text{PTHA0} + \text{PTHA1} * y + \text{PTHA2} * y^2$$

$$x = \text{pressure output} - \text{PTCA0} - \text{PTCA1} * t - \text{PTCA2} * t^2$$

$$n = x * \text{PTCB0} / (\text{PTCB0} + \text{PTCB1} * t + \text{PTCB2} * t^2)$$

$$\text{pressure (psia)} = \text{PA0} + \text{PA1} * n + \text{PA2} * n^2$$

