



SEA-BIRD
SCIENTIFIC

SBE Sea-Bird
Electronics

Sea-Bird Electronics
13431 NE 20th Street
Bellevue, Washington
98005 USA

Tel: +1 425-643-9866
seabird@seabird.com
www.seabird.com

SBE41-CP ALACE

Instrument Configuration

Instrument Serial Number: 41-8604
Instrument Firmware Version: V 7.2.5
Zero Conductivity Frequency: 2718.18
Communications Format: RS232
Communications Settings: 9600 baud, 8 Data Bits, No Parity

Installed Devices/Sensors

<i>Data Format</i>	<i>Measurement</i>	<i>Sensor Type</i>	<i>Serial Number</i>	<i>Rating</i>
Count	Temperature	Internal	N/A	N/A
Frequency	Conductivity	Internal	N/A	N/A
Count	Pressure	Druck	10374298	2000m(2000 dBar)

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 8604
CALIBRATION DATE: 15-Jul-16

SBE 41 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:


a0 = -8.240171e-004
a1 = 2.876289e-004
a2 = -3.496023e-006
a3 = 1.441044e-007

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	16258670.8	1.0000	0.0000
4.5000	13875818.5	4.5000	-0.0000
15.0000	8796544.3	15.0000	0.0000
18.5000	7603978.7	18.5000	-0.0000
23.9940	6085710.7	23.9940	-0.0000
29.0000	4998337.5	29.0000	-0.0000
32.5001	4370207.3	32.5001	0.0000

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature



Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 8604
 CALIBRATION DATE: 15-Jul-16

SBE 41 CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.797745e-001
 h = 1.332135e-001
 i = -3.363392e-004
 j = 4.220509e-005

CPcor = -9.5700e-008
 CTcor = 3.2500e-006
 WBOTC = -1.1955e-006


BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2718.18	0.00000	0.00000
1.0000	34.8461	2.97827	5462.75	2.97828	0.00001
4.5000	34.8262	3.28557	5670.28	3.28556	-0.00000
15.0000	34.7829	4.26795	6287.13	4.26794	-0.00001
18.5000	34.7737	4.61333	6489.77	4.61332	-0.00001
23.9940	34.7635	5.17102	6803.91	5.17103	0.00001
29.0000	34.7574	5.69376	7085.33	5.69376	0.00001
32.5001	34.7530	6.06621	7278.98	6.06620	-0.00001

$f = \text{Instrument Output(Hz)} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$

t = temperature (°C); p = pressure (decibars); $\delta = \text{CTcor}$; $\epsilon = \text{CPcor}$;

$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$

$\text{Residual (Siemens/meter)} = \text{instrument conductivity} - \text{bath conductivity}$



Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 8604

CALIBRATION DATE: 13-Jul-16

SBE 41 PRESSURE CALIBRATION DATA

2900 psia S/N 10374298

COEFFICIENTS:

PA0 = 5.468218e-001	PTCA0 = -3.669663e+003
PA1 = 3.929772e-004	PTCA1 = 9.202530e+001
PA2 = -2.899026e-013	PTCA2 = -1.337922e+000
PTHA0 = 2.893046e+002	PTCB0 = 2.507300e+001
PTHA1 = -6.021113e-005	PTCB1 = 1.200000e-003
PTHA2 = -1.059614e-012	PTCB2 = 0.000000e+000

PRESSURE SPAN CALIBRATION

THERMAL CORRECTION

PRESSURE (PSIA)	INSTRUMENT OUTPUT (counts)	THERMISTOR OUTPUT (counts)	COMPUTED PRESSURE (PSIA)	RESIDUAL (%FSR)	TEMP (°C)	THERMISTOR OUTPUT (counts)	INSTRUMENT OUTPUT (counts)
14.72	33965.1	4130994.8	14.77	0.00	32.50	3985520.00	35545.30
591.23	1504270.9	4128643.8	591.29	0.00	29.00	4036461.20	35564.42
1167.66	2977381.7	4127303.0	1167.64	-0.00	23.99	4109211.20	35476.66
1744.06	4453879.9	4126250.2	1744.06	-0.00	18.50	4188783.00	35229.43
2320.50	5933875.6	4125238.2	2320.57	0.00	15.00	4239445.40	35035.93
2896.82	7416310.1	4124237.2	2896.76	-0.00	4.50	4390760.80	34409.28
2320.45	5933630.9	4124339.0	2320.47	0.00	1.00	4441164.60	34079.60
1744.11	4454289.3	4124307.4	1744.20	0.00			
1167.64	2977151.5	4124377.2	1167.54	-0.00			
591.14	1503657.3	4124240.0	591.03	-0.00			
14.72	33858.8	4123794.4	14.73	0.00			

TEMPERATURE (°C)	SPAN (mV)
-5.00	25.07
35.00	25.11

y = thermistor output (counts)

$$t = PTHA0 + PTHA1 * y + PTHA2 * y^2$$

$$x = \text{instrument output} - PTCA0 - PTCA1 * t - PTCA2 * t^2$$

$$n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * t^2)$$

$$\text{pressure (PSIA)} = PA0 + PA1 * n + PA2 * n^2$$

$$\text{Residual (\%FSR)} = (\text{computed pressure} - \text{true pressure}) * 100 / \text{Full Scale Range}$$

Date, Offset (%FSR)

● 13-Jul-16 0.00

