



SEA-BIRD
SCIENTIFIC

SBE Sea-Bird
Electronics

Sea-Bird Electronics
13431 NE 20th Street
Bellevue, Washington
98005 USA

Tel: +1 425-643-9866
seabird@seabird.com
www.seabird.com

SBE41-CP ALACE

Instrument Configuration

Instrument Serial Number: 41-8761
Instrument Firmware Version: V 7.2.5
Zero Conductivity Frequency: 2569.09
Communications Format: RS232
Communications Settings: 9600 baud, 8 Data Bits, No Parity

Installed Devices/Sensors

<i>Data Format</i>	<i>Measurement</i>	<i>Sensor Type</i>	<i>Serial Number</i>	<i>Rating</i>
Count	Temperature	Internal	N/A	N/A
Frequency	Conductivity	Internal	N/A	N/A
Count	Pressure	Druck	10300915	2000m(2000 dBar)

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 8761
CALIBRATION DATE: 10-Aug-16

SBE 41 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:


a0 = -8.832531e-004
a1 = 3.016783e-004
a2 = -4.373507e-006
a3 = 1.617257e-007

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	15557476.5	1.0000	-0.0000
4.5000	13271231.1	4.5001	0.0001
15.0000	8402340.0	15.0000	-0.0000
18.5000	7260248.7	18.4999	-0.0001
23.9940	5807038.8	23.9941	0.0001
29.0000	4766972.7	29.0000	0.0000
32.5000	4166496.7	32.5000	-0.0000

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature



Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 8761
CALIBRATION DATE: 10-Aug-16

SBE 41 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.021110e+000
h = 1.555065e-001
i = -4.575802e-004
j = 5.721569e-005

CPcor = -9.5700e-008
CTcor = 3.2500e-006
WBOTC = 1.2973e-007


BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2569.09	0.00000	0.00000
1.0000	34.8805	2.98093	5086.89	2.98092	-0.00001
4.5000	34.8600	3.28844	5278.30	3.28845	0.00001
15.0000	34.8172	4.27171	5847.69	4.27171	-0.00001
18.5000	34.8081	4.61740	6034.85	4.61740	-0.00000
23.9940	34.7980	5.17558	6325.08	5.17559	0.00001
29.0000	34.7923	5.69883	6585.20	5.69882	-0.00001
32.5000	34.7881	6.07163	6764.24	6.07163	0.00000

$$f = \text{Instrument Output(Hz)} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$$

t = temperature (°C); p = pressure (decibars); δ = CTcor; ϵ = CPcor;

$$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$$

$$\text{Residual (Siemens/meter)} = \text{instrument conductivity} - \text{bath conductivity}$$



Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 8761
 CALIBRATION DATE: 02-Aug-16

SBE 41 PRESSURE CALIBRATION DATA
 2900 psia S/N 10300915

COEFFICIENTS:

PA0 = 4.862287e-001	PTCA0 = 9.710666e+003
PA1 = 3.934830e-004	PTCA1 = 5.261508e+001
PA2 = -3.063956e-013	PTCA2 = 1.535408e-001
PTHA0 = 3.107920e+002	PTCB0 = 2.514450e+001
PTHA1 = -6.022825e-005	PTCB1 = -1.000000e-004
PTHA2 = -1.432512e-012	PTCB2 = 0.000000e+000

PRESSURE SPAN CALIBRATION

THERMAL CORRECTION

PRESSURE (PSIA)	INSTRUMENT OUTPUT (counts)	THERMISTOR OUTPUT (counts)	COMPUTED PRESSURE (PSIA)	RESIDUAL (%FSR)	TEMP (°C)	THERMISTOR OUTPUT (counts)	INSTRUMENT OUTPUT (counts)
14.66	47105.7	4344903.4	14.72	0.00	32.50	4200897.80	49380.50
591.51	1514564.7	4341077.4	591.49	-0.00	29.00	4249249.00	49261.22
1168.17	2985296.5	4339973.0	1168.23	0.00	23.99	4318312.80	48991.47
1744.85	4459454.0	4338900.4	1744.98	0.00	18.50	4393878.60	48582.93
2321.62	5936966.8	4337859.6	2321.71	0.00	15.00	4441926.60	48321.10
2898.24	7417206.7	4336598.6	2898.16	-0.00	4.50	4585385.40	47803.03
2321.52	5936496.1	4336403.6	2321.53	0.00	1.00	4633103.00	47635.98
1744.80	4458864.9	4335788.0	1744.75	-0.00			
1167.91	2984272.1	4335824.2	1167.82	-0.00			
591.12	1513277.9	4335868.6	590.97	-0.00			
14.67	47176.9	4331254.4	14.72	0.00			
					TEMPERATURE (°C)	SPAN (mV)	
					-5.00	25.15	
					35.00	25.14	

y = thermistor output (counts)

$$t = PTHA0 + PTHA1 * y + PTHA2 * y^2$$

$$x = \text{instrument output} - PTCA0 - PTCA1 * t - PTCA2 * t^2$$

$$n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * t^2)$$

$$\text{pressure (PSIA)} = PA0 + PA1 * n + PA2 * n^2$$

$$\text{Residual (\%FSR)} = (\text{computed pressure} - \text{true pressure}) * 100 / \text{Full Scale Range}$$

Date, Offset (%FSR)

● 02-Aug-16 0.00

