



**SEA-BIRD**  
SCIENTIFIC

**SBE** Sea-Bird  
Electronics

Sea-Bird Electronics  
13431 NE 20<sup>th</sup> Street  
Bellevue, Washington  
98005 USA

Tel: +1 425-643-9866  
seabird@seabird.com  
www.seabird.com

## SBE41-CP ALACE

### Instrument Configuration

Instrument Serial Number: 41-9411  
Instrument Firmware Version: V 7.2.5  
Zero Conductivity Frequency: 2656.32  
Communications Format: RS232  
Communications Settings: 9600 baud, 8 Data Bits, No Parity

### Installed Devices/Sensors

| <i>Data Format</i> | <i>Measurement</i> | <i>Sensor Type</i> | <i>Serial Number</i> | <i>Rating</i>    |
|--------------------|--------------------|--------------------|----------------------|------------------|
| Count              | Temperature        | Internal           | N/A                  | N/A              |
| Frequency          | Conductivity       | Internal           | N/A                  | N/A              |
| Count              | Pressure           | Druck              | 10490656             | 2000m(2000 dBar) |

# Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9411  
CALIBRATION DATE: 27-Jan-17

SBE 41 TEMPERATURE CALIBRATION DATA  
ITS-90 TEMPERATURE SCALE

## COEFFICIENTS:


a0 = -8.046713e-004  
a1 = 2.879399e-004  
a2 = -3.513160e-006  
a3 = 1.439590e-007

| BATH TEMP<br>(° C) | INSTRUMENT<br>OUTPUT (counts) | INST TEMP<br>(° C) | RESIDUAL<br>(° C) |
|--------------------|-------------------------------|--------------------|-------------------|
| 1.0000             | 15221705.6                    | 1.0001             | 0.0001            |
| 4.5000             | 12984879.1                    | 4.4999             | -0.0001           |
| 14.9999            | 8220908.6                     | 15.0000            | 0.0001            |
| 18.5000            | 7103472.5                     | 18.5000            | 0.0000            |
| 23.9940            | 5681663.0                     | 23.9939            | -0.0001           |
| 29.0000            | 4664005.6                     | 28.9999            | -0.0001           |
| 32.5000            | 4076423.7                     | 32.5001            | 0.0001            |

n = Instrument Output (counts)

Temperature ITS-90 (°C) =  $1/\{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature



# Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9411  
CALIBRATION DATE: 27-Jan-17

SBE 41 CONDUCTIVITY CALIBRATION DATA  
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

## COEFFICIENTS:

g = -1.001712e+000  
h = 1.427275e-001  
i = -4.201310e-004  
j = 5.044519e-005

CPcor = -9.5700e-008  
CTcor = 3.2500e-006  
WBOTC = -8.4693e-007


| BATH TEMP<br>(° C) | BATH SAL<br>(PSU) | BATH COND<br>(S/m) | INSTRUMENT<br>OUTPUT (Hz) | INSTRUMENT<br>COND (S/m) | RESIDUAL<br>(S/m) |
|--------------------|-------------------|--------------------|---------------------------|--------------------------|-------------------|
| 22.0000            | 0.0000            | 0.00000            | 2656.32                   | 0.00000                  | 0.00000           |
| 1.0000             | 34.7735           | 2.97266            | 5291.98                   | 2.97266                  | 0.00001           |
| 4.5000             | 34.7538           | 3.27941            | 5491.95                   | 3.27940                  | -0.00001          |
| 14.9999            | 34.7121           | 4.26017            | 6086.64                   | 4.26017                  | 0.00000           |
| 18.5000            | 34.7034           | 4.60501            | 6282.08                   | 4.60501                  | -0.00000          |
| 23.9940            | 34.6943           | 5.16186            | 6585.13                   | 5.16187                  | 0.00001           |
| 29.0000            | 34.6901           | 5.68397            | 6856.74                   | 5.68396                  | -0.00001          |
| 32.5000            | 34.6885           | 6.05622            | 7043.79                   | 6.05622                  | 0.00000           |

$f = \text{Instrument Output(Hz)} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$

t = temperature (°C); p = pressure (decibars);  $\delta = \text{CTcor}$ ;  $\epsilon = \text{CPcor}$ ;

Conductivity (S/m) =  $(g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity



# Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9411  
CALIBRATION DATE: 31-Jan-17

SBE 41 PRESSURE CALIBRATION DATA  
2900 psia S/N 10490656

## COEFFICIENTS:

|         |                |         |                |
|---------|----------------|---------|----------------|
| PA0 =   | -2.578935e+000 | PTCA0 = | -1.004822e+005 |
| PA1 =   | 3.959532e-004  | PTCA1 = | -2.629601e+002 |
| PA2 =   | -3.114057e-013 | PTCA2 = | -1.621421e+000 |
| PTHA0 = | 3.258452e+002  | PTCB0 = | 2.459350e+001  |
| PTHA1 = | -6.278966e-005 | PTCB1 = | -1.900000e-003 |
| PTHA2 = | -1.313526e-012 | PTCB2 = | 0.000000e+000  |

## PRESSURE SPAN CALIBRATION

## THERMAL CORRECTION

| PRESSURE (PSIA) | INSTRUMENT OUTPUT (counts) | THERMISTOR OUTPUT (counts) | COMPUTED PRESSURE (PSIA) | RESIDUAL (%FSR) | TEMP (°C)        | THERMISTOR OUTPUT (counts) | INSTRUMENT OUTPUT (counts) |
|-----------------|----------------------------|----------------------------|--------------------------|-----------------|------------------|----------------------------|----------------------------|
| 14.72           | -63142.0                   | 4431496.8                  | 14.81                    | 0.00            | 32.50            | 4287302.00                 | -65015.70                  |
| 592.49          | 1395399.8                  | 4431024.2                  | 592.60                   | 0.00            | 29.00            | 4334609.20                 | -63691.20                  |
| 1170.29         | 2857060.4                  | 4430579.4                  | 1170.29                  | 0.00            | 23.99            | 4402000.80                 | -61911.70                  |
| 1747.99         | 4322277.2                  | 4430033.0                  | 1748.06                  | 0.00            | 18.50            | 4475751.20                 | -60116.90                  |
| 2325.79         | 5791046.5                  | 4429593.6                  | 2325.88                  | 0.00            | 15.00            | 4522691.80                 | -59090.80                  |
| 2903.60         | 7262759.1                  | 4429159.4                  | 2903.51                  | -0.00           | 4.50             | 4662875.00                 | -55961.50                  |
| 2325.83         | 5790815.4                  | 4429698.2                  | 2325.79                  | -0.00           | 1.00             | 4709612.00                 | -54954.50                  |
| 1748.27         | 4323352.4                  | 4430191.0                  | 1748.48                  | 0.01            |                  |                            |                            |
| 1170.34         | 2856621.1                  | 4430675.4                  | 1170.12                  | -0.01           | TEMPERATURE (°C) | SPAN (mV)                  |                            |
| 592.50          | 1394544.6                  | 4431228.6                  | 592.26                   | -0.01           | -5.00            | 24.60                      |                            |
| 14.72           | -63274.7                   | 4431659.8                  | 14.76                    | 0.00            | 35.00            | 24.53                      |                            |

y = thermistor output (counts)

$$t = PTHA0 + PTHA1 * y + PTHA2 * y^2$$

$$x = \text{instrument output} - PTCA0 - PTCA1 * t - PTCA2 * t^2$$

$$n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * t^2)$$

$$\text{pressure (PSIA)} = PA0 + PA1 * n + PA2 * n^2$$

$$\text{Residual (\%FSR)} = (\text{computed pressure} - \text{true pressure}) * 100 / \text{Full Scale Range}$$

Date, Offset (%FSR)

● 31-Jan-17 -0.00

