



SEA-BIRD
SCIENTIFIC

SBE Sea-Bird
Electronics

Sea-Bird Electronics
13431 NE 20th Street
Bellevue, Washington
98005 USA

Tel: +1 425-643-9866
seabird@seabird.com
www.seabird.com

SBE41-CP ALACE

Instrument Configuration

Instrument Serial Number: 41-9083
Instrument Firmware Version: V 7.2.5
Zero Conductivity Frequency: 2601.66
Communications Format: RS232
Communications Settings: 9600 baud, 8 Data Bits, No Parity

Installed Devices/Sensors

<i>Data Format</i>	<i>Measurement</i>	<i>Sensor Type</i>	<i>Serial Number</i>	<i>Rating</i>
Count	Temperature	Internal	N/A	N/A
Frequency	Conductivity	Internal	N/A	N/A
Count	Pressure	Kistler	4978230	2000m(2000 dBar)

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9083
CALIBRATION DATE: 21-Oct-16

SBE 41 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:


a0 = -6.744501e-004
a1 = 2.662417e-004
a2 = -2.141666e-006
a3 = 1.164305e-007

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	14113513.2	1.0001	0.0001
4.5000	12043041.4	4.4999	-0.0001
15.0000	7631130.0	15.0001	0.0001
18.5000	6595634.3	18.5001	0.0001
24.0001	5276336.2	24.0000	-0.0001
29.0000	4333810.0	28.9999	-0.0001
32.5001	3788660.7	32.5002	0.0001

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature



Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9083
CALIBRATION DATE: 21-Oct-16

SBE 41 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.949490e-001
h = 1.475053e-001
i = -3.102045e-004
j = 4.441151e-005

CPcor = -9.5700e-008
CTcor = 3.2500e-006
WBOTC = -1.4412e-006


BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2601.66	0.00000	0.00000
1.0000	34.8657	2.97978	5198.27	2.97978	-0.00000
4.5000	34.8459	3.28724	5395.04	3.28724	-0.00000
15.0000	34.8033	4.27019	5980.09	4.27020	0.00001
18.5000	34.7943	4.61577	6172.34	4.61577	0.00000
24.0001	34.7844	5.17442	6470.76	5.17441	-0.00001
29.0000	34.7793	5.69694	6737.61	5.69694	-0.00000
32.5001	34.7765	6.06985	6921.57	6.06985	0.00000

$$f = \text{Instrument Output(Hz)} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$$

t = temperature (°C); p = pressure (decibars); δ = CTcor; ϵ = CPcor;

$$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$$

$$\text{Residual (Siemens/meter)} = \text{instrument conductivity} - \text{bath conductivity}$$



Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9083
CALIBRATION DATE: 17-Oct-16

SBE 41 PRESSURE CALIBRATION DATA
2900 psia S/N 4978230

COEFFICIENTS:

PA0 =	-1.334002e+001	PTCA0 =	-1.232310e+004
PA1 =	3.914336e-004	PTCA1 =	-1.593413e+003
PA2 =	1.250018e-013	PTCA2 =	7.463460e-001
PTHA0 =	3.235879e+002	PTCB0 =	1.036264e+002
PTHA1 =	-9.279733e-005	PTCB1 =	7.695880e-003
PTHA2 =	3.078185e-012	PTCB2 =	0.000000e+000

PRESSURE SPAN CALIBRATION

THERMAL CORRECTION

PRESSURE (PSIA)	INSTRUMENT OUTPUT (counts)	THERMISTOR OUTPUT (counts)	COMPUTED PRESSURE (PSIA)	RESIDUAL (%FSR)	TEMP (°C)	THERMISTOR OUTPUT (counts)	INSTRUMENT OUTPUT (counts)
14.42	24890.0	3710523.0	14.54	0.00	32.50	3556269.80	10266.20
590.15	1497043.1	3708423.6	590.25	0.00	29.00	3605882.40	15679.16
1166.18	2968819.4	3707606.6	1166.29	0.00	24.00	3676953.00	23439.22
1742.20	4439255.2	3706820.4	1742.33	0.00	18.50	3755365.20	32020.41
2318.22	5908283.6	3705825.2	2318.37	0.01	15.00	3805993.60	37557.11
2894.17	7375173.8	3705070.0	2894.09	-0.00	4.50	3958086.20	54088.77
2318.20	5907851.5	3704916.0	2318.23	0.00	1.00	4009657.20	59674.83
1742.12	4437969.6	3704475.8	1741.91	-0.01			
1166.16	2967793.3	3704110.6	1166.01	-0.01	TEMPERATURE (°C)	SPAN (mV)	
589.95	1495536.0	3703785.8	589.84	-0.00	-7.31	103.57	
14.43	23551.0	3703240.2	14.33	-0.00	35.75	103.90	

y = thermistor output (counts)

$$t = PTHA0 + PTHA1 * y + PTHA2 * y^2$$

$$x = \text{instrument output} - PTCA0 - PTCA1 * t - PTCA2 * t^2$$

$$n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * t^2)$$

$$\text{pressure (PSIA)} = PA0 + PA1 * n + PA2 * n^2$$

$$\text{Residual (\%FSR)} = (\text{computed pressure} - \text{true pressure}) * 100 / \text{Full Scale Range}$$

Date, Offset (%FSR)

● 17-Oct-16 -0.00

