

**Product:** Oxygen Optode 4330

**Certificate No:** 1150421752520

**Serial No:** 2520

**Date:** 14.01.2016

---

This is to certify that this product has been pressure tested with the following instrument, and we confirm that no irregularities were found during the test:


Autoklav 800 bar – sn: 0210005

**Pressure readings:**

Pressure (Bar)	Pressure time (hour)
600	1

Date: 25 Jan 2016

Sign:



Vidar Selsvik, Production Engineer