

Certificate no: 4330_2944_00146790
Foil batch no: 1517M

Product: 4330
Calibration date: 12.04.2018

Serial no: 2944
Page 1 of 2

Index	Temperature reference(°C)	[O2] Reference(μM)	Temperature raw data(mV)	Phase reading(°)
0	30.195	1.84	-54.807	59.37
1	20.092	1.50	269.027	60.38
2	10.245	1.48	581.893	61.10
3	0.831	1.84	855.987	61.70
4	0.877	19.96	854.747	59.24
5	0.919	41.92	853.587	56.57
6	0.950	64.32	852.767	54.13
7	0.980	104.55	851.953	50.35
8	1.017	150.41	850.947	46.77
9	1.055	217.74	849.907	42.57
10	1.113	325.76	848.327	37.57
11	1.144	425.27	847.487	34.18
12	1.152	535.16	847.287	31.33
13	10.671	19.34	568.733	57.69
14	10.491	33.93	574.280	55.19
15	10.411	51.93	576.720	52.48
16	10.343	85.64	578.840	48.21
17	10.297	117.99	580.267	44.89
18	10.272	171.47	581.040	40.53
19	10.245	254.41	581.840	35.63
20	10.212	340.06	582.867	32.06
21	10.184	427.65	583.740	29.38
22	20.474	16.49	256.760	56.39
23	20.383	28.73	259.653	53.53
24	20.324	41.75	261.560	50.88
25	20.269	68.70	263.347	46.33
26	20.235	94.58	264.400	42.85
27	20.228	137.58	264.600	38.40
28	20.222	203.07	264.800	33.59
29	20.220	272.52	264.900	30.08
30	20.218	340.77	264.940	27.57
31	30.165	14.28	-53.807	55.07
32	30.175	23.63	-54.140	52.18
33	30.200	34.51	-54.920	49.30
34	30.227	56.40	-55.760	44.58
35	30.255	78.87	-56.640	40.82
36	30.275	113.00	-57.273	36.48
37	30.290	168.27	-57.753	31.67
38	30.306	225.58	-58.253	28.32
39	30.320	281.89	-58.673	25.95


Certificate no: 4330_2944_00146790
Foil batch no: 1517M

Product: 4330
Calibration date: 12.04.2018


Serial no: 2944
Page 2 of 2

Index	0	1	2	3	4	5	6
SVUFoilCoef	2.73184E-03	1.19010E-04	2.20741E-06	1.75959E02	-2.06434E-01	-4.15203E01	3.47464E00
TempCoef	2.84510E01	-3.16369E-02	3.41817E-06	-4.84951E-09	0.00000E00	0.00000E00	

Oxygen validation



Temperature validation



Date:12.04.2018

Tor-Ove Kvalvaag
Tor-Ove Kvalvaag, Calibration Engineer