



**SEA-BIRD**  
SCIENTIFIC

## SBE41-CP ALACE

### Instrument Configuration

Instrument Serial Number: 41-11714  
Instrument Firmware Version: V 7.2.5  
Zero Conductivity Frequency: 2507.06  
Communications Format: RS232  
Communications Settings: 9600 baud, 8 Data Bits, No Parity

### Installed Devices/Sensors

<i>Data Format</i>	<i>Measurement</i>	<i>Sensor Type</i>	<i>Serial Number</i>	<i>Rating</i>
Count	Temperature	Internal	N/A	N/A
Frequency	Conductivity	Internal	N/A	N/A
Count	Pressure	Druck	11162139	2000m(2000 dBar)



Sea-Bird Scientific  
13431 NE 20<sup>th</sup> Street  
Bellevue, WA 98005  
USA

+1 425-643-9866  
seabird@seabird.com  
www.seabird.com

SENSOR SERIAL NUMBER: 11714  
CALIBRATION DATE: 29-Mar-19

SBE 41 TEMPERATURE CALIBRATION DATA  
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:


a0 = -8.154169e-004  
a1 = 2.882167e-004  
a2 = -3.382606e-006  
a3 = 1.449994e-007

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0005	13552274.0	1.0006	0.0001
4.5000	11589542.9	4.4999	-0.0001
15.0000	7389551.8	15.0001	0.0001
18.5000	6399529.9	18.5000	0.0000
23.9940	5136232.3	23.9939	-0.0001
29.0000	4229038.4	28.9999	-0.0001
32.5000	3703877.1	32.5001	0.0001

n = Instrument Output (counts)

Temperature ITS-90 (°C) =  $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature





Sea-Bird Scientific  
13431 NE 20<sup>th</sup> Street  
Bellevue, WA 98005  
USA

+1 425-643-9866  
seabird@seabird.com  
www.seabird.com

SENSOR SERIAL NUMBER: 11714  
CALIBRATION DATE: 29-Mar-19

SBE 41 CONDUCTIVITY CALIBRATION DATA  
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.011250e+000  
h = 1.615161e-001  
i = -3.899231e-004  
j = 5.633798e-005

CPcor = -9.5700e-008  
CTcor = 3.2500e-006  
WBOTC = -6.3148e-007


BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2507.06	0.00000	0.00000
1.0005	34.7798	2.97319	4975.20	2.97319	0.00000
4.5000	34.7602	3.27995	5162.66	3.27995	-0.00000
15.0000	34.7179	4.26082	5720.24	4.26082	0.00000
18.5000	34.7089	4.60566	5903.49	4.60565	-0.00001
23.9940	34.6988	5.16246	6187.67	5.16247	0.00001
29.0000	34.6933	5.68444	6442.35	5.68442	-0.00001
32.5000	34.6898	6.05642	6617.69	6.05643	0.00000

$f = \text{Instrument Output(Hz)} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$

t = temperature (°C); p = pressure (decibars);  $\delta = \text{CTcor}$ ;  $\epsilon = \text{CPcor}$ ;

$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity





Sea-Bird Scientific  
13431 NE 20<sup>th</sup> Street  
Bellevue, WA 98005  
USA

+1 425-643-9866  
seabird@seabird.com  
www.seabird.com

SENSOR SERIAL NUMBER: 11714  
CALIBRATION DATE: 20-Mar-19

SBE 41 PRESSURE CALIBRATION DATA  
2900 psia S/N 11162139

COEFFICIENTS:

PA0 =	2.648949e-001	PTCA0 =	-2.290742e+003
PA1 =	3.935898e-004	PTCA1 =	3.845018e+001
PA2 =	-2.899420e-013	PTCA2 =	-4.099843e-001
PTHA0 =	2.967651e+002	PTCB0 =	3.097977e+005
PTHA1 =	-6.237470e-005	PTCB1 =	1.397689e+001
PTHA2 =	-8.532745e-013	PTCB2 =	6.792308e-002

PRESSURE SPAN CALIBRATION

THERMAL CORRECTION

PRESSURE (PSIA)	INSTRUMENT OUTPUT (counts)	THERMISTOR OUTPUT (counts)	COMPUTED PRESSURE (PSIA)	RESIDUAL (%FSR)	TEMP (°C)	THERMISTOR OUTPUT (counts)	INSTRUMENT OUTPUT (counts)
14.55	34654.0	4148395.4	14.52	-0.00	32.50	4016114.60	36680.50
591.43	1503908.3	4148297.2	591.47	0.00	29.00	4066569.40	36675.44
1168.40	2976112.0	4147790.6	1168.33	-0.00	23.99	4138774.40	36617.90
1745.32	4451943.4	4147362.4	1745.34	0.00	18.50	4217849.80	36451.54
2322.24	5930990.2	4146959.4	2322.35	0.00	15.00	4268173.60	36332.86
2899.18	7412578.8	4146741.4	2899.08	-0.00	4.50	4418354.60	36075.47
2322.16	5930661.1	4147436.4	2322.23	0.00	1.00	4468710.20	35920.72
1745.41	4452025.6	4147946.4	1745.38	-0.00	TEMPERATURE (°C)      SPAN		
1168.27	2975730.5	4148484.2	1168.18	-0.00			
591.28	1503413.5	4149023.4	591.28	0.00			
14.55	34837.5	4149641.8	14.60	0.00			
					2.12	309827.70	
					20.82	310118.15	
					32.81	310329.42	

y = thermistor output (counts)

t = PTHA0 + PTHA1 \* y + PTHA2 \* y<sup>2</sup>

x = instrument output - PTCA0 - PTCA1 \* t - PTCA2 \* t<sup>2</sup>

n = x \* PTCB0 / (PTCB0 + PTCB1 \* t + PTCB2 \* t<sup>2</sup>)

pressure (PSIA) = PA0 + PA1 \* n + PA2 \* n<sup>2</sup>

Residual (%FSR) = (computed pressure - true pressure) \* 100 / Full Scale Range

Date, Offset (%FSR)

● 20-Mar-19 -0.00

