



**SEA-BIRD**  
SCIENTIFIC

## SBE41-CP ALACE

### Instrument Configuration

Instrument Serial Number: 41-12964  
Instrument Firmware Version: 7.2.5  
Zero Conductivity Frequency: 2695.17  
Communications Format: RS232  
Communications Settings: 9600 baud, 8 Data Bits, No Parity

### Installed Devices/Sensors

<i>Data Format</i>	<i>Measurement</i>	<i>Sensor Type</i>	<i>Serial Number</i>	<i>Rating</i>
Count	Temperature	Internal	N/A	N/A
Frequency	Conductivity	Internal	N/A	N/A
Count	Pressure	Druck	11499790	2000m(2000 dBar)



Sea-Bird Scientific  
13431 NE 20<sup>th</sup> Street  
Bellevue, WA 98005  
USA

+1 425-643-9866  
seabird@seabird.com  
www.seabird.com

SENSOR SERIAL NUMBER: 12964  
CALIBRATION DATE: 20-Jun-20

SBE 41 TEMPERATURE CALIBRATION DATA  
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:


a0 = -8.536876e-004  
a1 = 2.965449e-004  
a2 = -3.912384e-006  
a3 = 1.556523e-007

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	13424048.4	1.0000	0.0000
4.5000	11476333.0	4.5000	-0.0000
15.0000	7311579.1	15.0000	0.0000
18.5000	6330404.3	18.5000	-0.0000
24.0000	5077601.8	24.0000	0.0000
28.9999	4180414.1	28.9999	-0.0000
32.5000	3660495.6	32.5000	0.0000

n = Instrument Output (counts)

Temperature ITS-90 (°C) =  $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature





Sea-Bird Scientific  
13431 NE 20<sup>th</sup> Street  
Bellevue, WA 98005  
USA

+1 425-643-9866  
seabird@seabird.com  
www.seabird.com

SENSOR SERIAL NUMBER: 12964  
CALIBRATION DATE: 20-Jun-20

SBE 41 CONDUCTIVITY CALIBRATION DATA  
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.011468e+000  
h = 1.396080e-001  
i = -2.354497e-004  
j = 3.682894e-005

CPcor = -9.5700e-008  
CTcor = 3.2500e-006  
WBOTC = 1.5149e-006


BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2695.17	0.00000	0.00000
1.0000	34.7401	2.97007	5344.32	2.97008	0.00001
4.5000	34.7218	3.27669	5545.62	3.27668	-0.00001
15.0000	34.6829	4.25698	6144.36	4.25697	-0.00001
18.5000	34.6750	4.60164	6341.18	4.60164	0.00000
24.0000	34.6673	5.15891	6646.77	5.15891	0.00001
28.9999	34.6644	5.68022	6920.09	5.68023	0.00000
32.5000	34.6633	6.05232	7108.55	6.05232	-0.00001

$f = \text{Instrument Output(Hz)} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$

t = temperature (°C); p = pressure (decibars);  $\delta = \text{CTcor}$ ;  $\epsilon = \text{CPcor}$ ;

$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity





Sea-Bird Scientific  
13431 NE 20<sup>th</sup> Street  
Bellevue, WA 98005  
USA

+1 425-643-9866  
seabird@seabird.com  
www.seabird.com

SENSOR SERIAL NUMBER: 12964  
CALIBRATION DATE: 15-Jun-20

SBE 41 PRESSURE CALIBRATION DATA  
2900 psia S/N 11499790

#### COEFFICIENTS:

PA0 =	2.487153e-001	PTCA0 =	-2.651459e+003
PA1 =	3.914315e-004	PTCA1 =	5.374846e+001
PA2 =	-2.872817e-013	PTCA2 =	-1.129533e+000
PTHA0 =	3.211909e+002	PTCB0 =	3.138039e+005
PTHA1 =	-6.170848e-005	PTCB1 =	2.071003e+000
PTHA2 =	-1.390796e-012	PTCB2 =	8.556573e-002

#### PRESSURE SPAN CALIBRATION

#### THERMAL CORRECTION

PRESSURE (PSIA)	INSTRUMENT OUTPUT (counts)	THERMISTOR OUTPUT (counts)	COMPUTED PRESSURE (PSIA)	RESIDUAL (%FSR)	TEMP (°C)	THERMISTOR OUTPUT (counts)	INSTRUMENT OUTPUT (counts)
14.44	34374.5	4410839.2	14.49	0.00	32.50	4267811.60	35688.50
590.93	1508602.8	4404780.0	590.73	-0.01	29.00	4315292.00	35857.76
1167.65	2988035.0	4403025.0	1167.75	0.00	24.00	4383049.80	35931.85
1744.44	4470135.2	4401743.6	1744.54	0.00	18.50	4457392.80	35803.09
2321.18	5955202.4	4400527.6	2321.22	0.00	15.00	4504583.80	35668.64
2897.86	7443376.8	4399410.6	2897.84	-0.00	4.50	4645616.20	35421.18
2321.07	5954648.3	4399296.4	2321.01	-0.00	1.00	4692496.40	35274.31
1744.60	4470254.5	4399135.4	1744.58	-0.00	<div>TEMPERATURE (°C)      SPAN</div> <div>1.18                      313806.51</div> <div>20.45                     313882.07</div> <div>33.57                     313969.87</div>		
1167.34	2986935.3	4399029.2	1167.31	-0.00			
593.75	1523946.2	4398995.2	596.71	0.10			
14.44	34361.8	4397065.4	14.48	0.00			

y = thermistor output (counts)

t = PTHA0 + PTHA1 \* y + PTHA2 \* y<sup>2</sup>

x = instrument output - PTCA0 - PTCA1 \* t - PTCA2 \* t<sup>2</sup>

n = x \* PTCB0 / (PTCB0 + PTCB1 \* t + PTCB2 \* t<sup>2</sup>)

pressure (PSIA) = PA0 + PA1 \* n + PA2 \* n<sup>2</sup>

Residual (%FSR) = (computed pressure - true pressure) \* 100 / Full Scale Range

Date, Offset (%FSR)

● 15-Jun-20 0.00

