



SEA-BIRD
ANALOG DATA

Sea-Bird Electronics, Inc.
 13431 NE 20th Street
 Bellevue, WA 98005 United States

Phone
 Fax

+1-425-643-9866
 +1-425-643-9954
 www.seabird.com

Service Request
 Date

17-AUG-2019

Sales Order

315682035

SERVICE REPORT

PRODUCT INFORMATION

Item: SLOCUM.LEGACY
 Item Description: (LEGACY) Slocum Glider
 Serial: 9102

Special Notes

Services Requested:
 Evaluate/Repair Instrumentation.
 Perform Routine Calibration Service.
 Replace Antifoulant Device(s).
 Services Performed:
 Perform initial diagnostic evaluation.
 Performed pressure calibration.
 Performed "POST" cruise calibration.
 Replaced the lithium back-up battery(s).
 Installed NEW AF24173 Anti-foulant cylinder(s).

Item	Item Description	Qty
CAL_SLOCUM	Calibrate SLOCUM conductivity and temperature sensors	1
CNCRTSLOCUM	Confirm & Re-certify Webb SLOCUM Glider CTD	1
REPLACEAF	Extra charge to install one antifoulant device, includes one 801542.1.	1
PCAL_SLOCUM	Calibrate SLOCUM pressure sensor	1

Unbilled Items

Item	Item Description	Qty
801542.1	AF24173 ANTI-FOULANT, SINGLE CYLINDER, V2	1
22096	LITHIUM COIN BATTERY, WITH TABS, BR1632A/HA	1



SEA-BIRD
SCIENTIFIC

Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 9102
CALIBRATION DATE: 12-Jul-19

Slocum Payload CTD TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:


a0 = -3.598741e-005
a1 = 2.905977e-004
a2 = -3.222216e-006
a3 = 1.711379e-007

BATH TEMP (°C)	INSTRUMENT OUTPUT (counts)	INST TEMP (°C)	RESIDUAL (°C)
1.0000	569620.0	1.0000	0.0000
4.5000	487360.0	4.5000	-0.0000
15.0000	311267.9	15.0000	0.0000
18.5000	269728.2	18.5000	0.0000
24.0000	216648.7	24.0000	-0.0000
29.0000	178597.7	28.9999	-0.0001
32.5000	156529.4	32.5001	0.0001

n = Instrument Output (counts)

$$\text{Temperature ITS-90 (°C)} = 1/\{a0 + a1[I(n)] + a2[I(n)^2] + a3[I(n)^3]\} - 273.15$$

Residual (°C) = instrument temperature - bath temperature





SEA-BIRD
SCIENTIFIC

Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 9102
CALIBRATION DATE: 12-Jul-19

Slocum Payload CTD CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35, 15.0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.842208e-001
h = 1.363516e-001
i = -2.630428e-004
j = 3.741392e-005

CPcor = -9.5700e-008
CTcor = 3.2500e-006
WBOTC = -1.2386e-007

BATH TEMP (°C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2691.00	0.00000	0.00000
1.0000	34.7040	2.96728	5389.88	2.96728	0.00001
4.5000	34.6845	3.27351	5594.23	3.27350	-0.00001
15.0000	34.6414	4.25242	6201.72	4.25242	0.00000
18.5000	34.6321	4.59656	6401.31	4.59656	-0.00000
24.0000	34.6220	5.15291	6711.12	5.15291	-0.00000
29.0000	34.6160	5.67319	6988.09	5.67319	0.00000
32.5000	34.6118	6.04435	7178.94	6.04435	-0.00000

f = Instrument Output(Hz) * sqrt(1.0 + WBOTC * t) / 1000.0


t = temperature (°C); p = pressure (decibars); δ = CTcor; e = CPcor;

Conductivity (S/m) = (g + h * f² + i * f + j * f³) / (1 + δ * t + e * p)

Residual (Siemens/meter) = instrument conductivity - bath conductivity

Date, Slope Correction

- 19-Jul-13 0.9995673
- ▲ 12-Jul-19 1.0000000



**POST CRUISE
CALIBRATION**



SEA-BIRD
SCIENTIFIC

Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 9102
CALIBRATION DATE: 11-Jul-19

Slocum Payload CTD PRESSURE CALIBRATION DATA
1450 psia S/N 3912712

COEFFICIENTS:

PA0 = 2.195155e-001
PA1 = 4.530344e-003
PA2 = -1.960528e-011
PTempa0 = -7.046686e+001
PTempa1 = 5.229587e-002
PTempa2 = -6.773806e-007

PTCA0 = 5.248726e+005
PTCA1 = 1.470111e+000
PTCA2 = 1.759149e-002
PTCB0 = 2.534875e+001
PTCB1 = -1.650000e-003
PTCB2 = 0.000000e+000

PRESSURE SPAN CALIBRATION

PRESSURE (PSIA)	INSTRUMENT OUTPUT (counts)	THERMISTOR OUTPUT (volts)	COMPUTED PRESSURE (PSIA)	RESIDUAL (%FSR)	TEMP (°C)	THERMISTOR OUTPUT (volts)	INSTRUMENT OUTPUT (counts)
14.68	528107.0	1829.0	14.70	0.00	32.50	2022	528158.50
314.94	594280.0	1830.0	314.84	-0.01	29.00	1952	528150.10
614.89	660459.0	1831.0	614.84	-0.00	24.00	1851	528141.60
914.94	726682.0	1835.0	914.87	-0.00	18.50	1740	528125.90
1214.90	792940.0	1836.0	1214.89	-0.00	15.00	1671	528116.30
1464.88	848170.0	1836.0	1464.84	-0.00	4.50	1461	528101.40
1214.84	792947.0	1836.0	1214.92	0.01	1.00	1392	528094.00
914.89	726699.0	1835.0	914.95	0.00			
614.86	660476.0	1835.0	614.92	0.00			
314.84	594288.0	1835.0	314.88	0.00			
14.68	528105.0	1836.0	14.69	0.00			
					TEMPERATURE (°C)		SPAN
					-5.00		25.36
					35.00		25.29

THERMAL CORRECTION

$$y = \text{thermistor output (counts)}$$

$$t = \text{PTempaPA0} + \text{PTempa1} * y + \text{PTempa2} * y^2$$

$$x = \text{instrument output} - \text{PTCA0} - \text{PTCA1} * t - \text{PTCA2} * t^2$$

$$n = x * \text{PTCB0} / (\text{PTCB0} + \text{PTCB1} * t + \text{PTCB2} * t^2)$$

$$\text{pressure (PSIA)} = \text{PA0} + \text{PA1} * n + \text{PA2} * n^2$$

$$\text{Residual (\%FSR)} = (\text{computed pressure} - \text{true pressure}) * 100 / \text{Full Scale Range}$$

