



Sea-Bird Scientific
 13431 NE 20th Street
 Bellevue, WA 98005
 USA

+1 425-643-9866
 seabird@seabird.com
 www.seabird.com

SENSOR SERIAL NUMBER: 0026
 CALIBRATION DATE: 25-Oct-18

Glider Payload CTD CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.940420e-001 CPcor = -9.5700e-008
 h = 1.619515e-001 CTcor = 3.2500e-006
 i = -4.104616e-004 WBOTC = 8.3917e-007
 j = 5.785648e-005

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2482.54	0.00000	0.00000
1.0000	34.6685	2.96453	4953.38	2.96453	-0.00000
4.4999	34.6489	3.27047	5140.73	3.27048	0.00000
14.9999	34.6073	4.24867	5697.83	4.24867	0.00000
18.5000	34.5985	4.59258	5880.91	4.59259	0.00001
24.0000	34.5891	5.14855	6165.07	5.14854	-0.00001
29.0000	34.5835	5.66846	6419.11	5.66847	0.00000
32.5000	34.5795	6.03935	6594.16	6.03935	-0.00000

$$f = \text{Instrument Output(Hz)} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$$

t = temperature (°C); p = pressure (decibars); δ = CTcor; ε = CPcor;

$$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

