



OGS (ISTITUTO NAZIONALE DI OCEANOGRAFIA E DI GEOFISICA SPERIMENTALE)  
SEZIONE DI OCEANOGRAFIA

## Activity Report

### **Post-deployment evaluation and recalibration of temperature and conductivity sensors on the ARVOR float CTDs**

SBE 41 ALACE-CP-MO V 1.3, Serial No. 2754  
&

SBE 41 ALACE-CP-MO V 1.2, Serial No. 4522

**Rajesh Nair, Nevio Medeot**  
Oceanographic Calibration Centre, Oceanography Section

Approved for release by:

Dr. Paola Del Negro  
The Director, Oceanography Section.

A handwritten signature in black ink, appearing to read 'P. Del Negro'.

**REL. 2014/76 OCE 28 CTO-TECDEV, Borgo Grotta Gigante (TS), 09 December 2014**



## Section A

### Background information

#### A1. Instrument description

The SBE 41/41CP CTD module, designed for use on profiling floats typically during the ascent phase, samples temperature (T), conductivity (C), and pressure continuously at a rate of 1 Hz when powered on. The module is manufactured by Sea-Bird Electronics, Inc., and incorporates their trademark pump-controlled, T-C ducted flow configuration to minimize salinity spiking caused by mismatch of temperature and conductivity measurements. The pump is normally turned off between profiles. The indispensable anti-fouling protection necessary to ensure high-quality data includes anti-foulant devices to minimize bio-fouling of the module's conductivity cell, a U-shaped flow path, and a programmable pump cutoff setting, (usually 10-5 dbars) to switch off the integral pump as the float approaches the sea surface. The U-shaped flow path prevents oils and contaminants from being ingested as the float breaks through the surface skin of the sea. It also prevents water from flowing through the system due to waves or currents while the float is at the surface transmitting the data it has collected.

The characteristics of the T and C sensors when these CTD modules are mounted on ARVOR-C profiling floats manufactured by nke Instrumentation are reported in the table below.

**Table.** ARVOR-C temperature and conductivity sensor specifications.

PARAMETER	SPECIFICATION	
	Initial Accuracy	Resolution
Temperature (°C, ITS-90)	±0.002	0.001
Conductivity (equivalent salinity, PSS-78)	±0.005	0.001

#### A2. Units under test

Instrument	Vector	Launch date & location	Recovery date & location	Period of deployment	Area of Operation
SBE 41 ALACE-CP-MO, V 1.3, Serial No. 2754	nke ARVOR-C: @BT2011 04 70, O/N 11IT ARC 01	10/10/2011, Tyrrhenian Sea	27/10/2011, Tyrrhenian Sea	~17 days	Western Mediterranean Sea
		18/10/2012, Adriatic Sea	27/10/2012, Adriatic Sea	~4 days	Eastern Mediterranean Sea

Instrument	Vector	Launch date & location	Recovery date & location	Period of deployment	Area of Operation
SBE 41 ALACE-CP-MO, V 1.2, Serial No. 4522	nke ARVOR-I: @BT2012 03 14, O/N 12IT-ARI-04	03/05/2012, Tyrrhenian Sea	22/04/2014, Tarragona (Spain)	~2 years	Western Mediterranean Sea



## Section B

### Calibration equipment & pre-treatment of sensors

#### B1. Calibration apparatus

Instrument description	Model	Serial number
Laboratory Salinometer <sup>a</sup>	Guildline Autosal 8400B	65744
Standard Platinum Resistance Thermometer (SPRT) <sup>b</sup>	Hart 5699	0103
Precision Digital Thermometer	Hart - Fluke 1595A	B46159
Seawater Calibration Bath <sup>c</sup>	Hart 7052	A1A003
PC Interface Box for SBE 41CP CTD Module	FL-CTD2008	OGS001
DC Power Supply	Agilent E3631A	MY40007143

<sup>a</sup>Laboratory Salinometer standardized using freshly-opened IAPSO Standard Seawater bottles (Batch: P156) at the start of every 24 hours of operation;

<sup>b</sup>Reference SPRT calibration to sub-range 11 (0.01 °C - 29.7646 °C) of the International Temperature Scale of 1990 (ITS-90) last checked on 25 June 2014 using a Hart Scientific 5901 Triple Point of Water cell and a Hart Scientific 5943 Gallium Melting Point cell;

<sup>c</sup>Calibration bath filled with natural, filtered seawater (filter size/type: 0.22 µm/Millipore).

#### B2. Pre-treatment of sensors

The T and C sensors of the two SBE 41 ALACE-CP units were rinsed by flushing fresh deionized water through their U-shaped sample flow paths for a few minutes for the purpose of the post-deployment evaluation. No other cleaning operation was performed.

On the contrary, preparing both units for calibration involved repeated cleaning of their C sensors, implementing the manufacturer's instructions for this operation ("Instructions for Care and Cleaning of Conductivity Cells", Sea-Bird Electronics, Inc. Application Note No. 2D, last revision: March 2014). After each cleaning cycle, the C sensor responses were tested at a convenient bath temperature set-point (21 °C for unit 2754 and 20 °C for unit 4522) until the conductivity residuals from consecutive sets of measurements showed no further significant differences from the corresponding reference values.



# Section C

## Data & Results

# **POST-DEPLOYMENT EVALUATION**



## TEMPERATURE

Test date: 13 November 2014

Ambient conditions:

Temperature:  $21.7^{\circ}\text{C} \pm 1^{\circ}\text{C}$

Relative Humidity:  $52\% \pm 10\%$

Atmospheric pressure: 978.4 hPa

### As-received temperature calibration coefficients

$$a_0 = 1.574572e-05$$

$$a_1 = 2.773593e-04$$

$$a_2 = -2.712031e-06$$

$$a_3 = 1.570812e-07$$

$$T (\text{°C}) = 1/ \{ [a_0 + a_1 [\ln(n)] + a_2 [\ln^2(n)] + a_3 [\ln^3(n)]] \} - 273.15$$

REF TEMP (°C)	INST OUTPUT (n)	INST TEMP (°C)	TEMP RESIDUAL (°C)
3.0327	654242.0	3.0316	-0.0011
5.1475	594806.6	5.1464	-0.0011
10.2503	475043.1	10.2489	-0.0014
15.2259	384041.5	15.2237	-0.0022
20.1292	313312.1	20.1275	-0.0017
25.0381	257042.9	25.0362	-0.0019
27.0100	237771.2	27.0079	-0.0021

where:

**REF TEMP** = bath set-point temperature (°C, ITS-90), measured using the reference Standard Platinum Resistance Thermometer;

**INST OUTPUT** = Instrument output at **REF TEMP**;

**INST TEMP** = bath temperature (°C, ITS-90), measured by Instrument;

**TEMP RESIDUAL** = temperature residual: **INST TEMP** - **REF TEMP**.



## CONDUCTIVITY

Test date: 13 November 2014

Ambient conditions:

Temperature:  $21.7^{\circ}\text{C} \pm 1^{\circ}\text{C}$

Relative Humidity:  $52\% \pm 10\%$

Atmospheric pressure: 978.4 hPa

### As-received conductivity calibration coefficients

$$\begin{aligned} g &= -1.010586e+00 & CPcor &= -9.570001e-08 \\ h &= 1.483483e-01 & CTcor &= 3.250000e-06 \\ i &= -3.440286e-04 & WBOTC &= 1.593064e-07 \\ j &= 4.595743e-05 \end{aligned}$$

$$\text{Conductivity (Siemens/m)} = (g + hf^2 + if^3 + jf^4) / (1 + \delta t + \varepsilon p)$$

**Note:**  $f = \text{Inst Freq} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$

$t$  = Temperature ( $^{\circ}\text{C}$ );  $p$  = Pressure (dBar);  $\delta$  = CTcor;  $\varepsilon$  = CPcor.

REF TEMP ( $^{\circ}\text{C}$ )	BATH SAL (PSS-78)	REF COND (S/m)	INST OUTPUT (Hz)	INST COND (S/m)	COND RESIDUAL (S/m)	SAL RESIDUAL (PSS-78)
21.1527	0	0	2615.24 <sup>1</sup>	0.00004	0.00004	-
3.0327	37.5471	3.37605	5447.25	3.37612	0.00007	0.0021
5.1475	37.5479	3.57801	5570.83	3.57801	0.00000	0.0013
10.2503	37.5511	4.08188	5867.65	4.08180	-0.00008	0.0007
15.2259	37.5531	4.59365	6154.22	4.59355	-0.00010	0.0012
20.1292	37.5542	5.11560	6433.18	5.11574	0.00014	0.0027
25.0381	37.5587	5.65416	6708.41	5.65427	0.00011	0.0024
27.0100	37.5701	5.87580	6818.37	5.87594	0.00014	0.0028

where:

**REF TEMP** = bath set-point temperature ( $^{\circ}\text{C}$ , ITS-90), measured using the reference Standard Platinum Resistance Thermometer;

**BATH SAL** = bath salinity (PSS-78), measured using the reference Laboratory Salinometer;

**REF COND** = conductivity set-point (S/m), obtained from inverted **BATH SAL**;

**INST OUTPUT** = Instrument output frequency (Hz) at **REF TEMP**;

**INST COND** = bath conductivity (S/m), measured by Instrument;

**COND RESIDUAL** = conductivity residual: **INST COND** - **REF COND**;

**SAL RESIDUAL** = salinity residual: **INST SAL** - **REF SAL**.

<sup>1</sup> In air, with a perfectly dry conductivity cell.



## TEMPERATURE

Test date: 21 November 2014

Ambient conditions:

Temperature:  $21.2^{\circ}\text{C} \pm 1^{\circ}\text{C}$

Relative Humidity:  $47\% \pm 10\%$

Atmospheric pressure: 981.4 hPa

### As-received temperature calibration coefficients

$$a_0 = -1.098629e-06$$

$$a_1 = 2.728035e-04$$

$$a_2 = -2.385421e-06$$

$$a_3 = 1.468875e-07$$

$$T (\text{°C}) = 1 / \{ [a_0 + a_1 [\ln(n)] + a_2 [\ln^2(n)] + a_3 [\ln^3(n)]] \} - 273.15$$

REF TEMP (°C)	INST OUTPUT (n)	INST TEMP (°C)	TEMP RESIDUAL (°C)
3.0326	761372.9	3.0328	0.0002
5.1484	692136.4	5.1482	-0.0002
10.2509	552686.2	10.2508	-0.0001
15.2254	446755.3	15.2249	-0.0005
20.1297	364410.1	20.1294	-0.0003
25.0393	298915.7	25.0384	-0.0009
27.0118	276483.5	27.0106	-0.0012

where:

**REF TEMP** = bath set-point temperature (°C, ITS-90), measured using the reference Standard Platinum Resistance Thermometer;

**INST OUTPUT** = Instrument output at **REF TEMP**;

**INST TEMP** = bath temperature (°C, ITS-90), measured by Instrument;

**TEMP RESIDUAL** = temperature residual: **INST TEMP** - **REF TEMP**.



## CONDUCTIVITY

Test date: 21 November 2014

Ambient conditions:

Temperature:  $21.2^{\circ}\text{C} \pm 1^{\circ}\text{C}$

Relative Humidity:  $47\% \pm 10\%$

Atmospheric pressure: 981.4 hPa

### As-received conductivity calibration coefficients

$$\begin{aligned} g &= -9.861773\text{e-01} & \text{CPcor} &= -9.570001\text{e-08} \\ h &= 1.523203\text{e-01} & \text{CTcor} &= 3.250000\text{e-06} \\ i &= -4.120115\text{e-04} & \text{WBOTC} &= -2.352400\text{e-07} \\ j &= 5.421455\text{e-05} \end{aligned}$$

$$\text{Conductivity (Siemens/m)} = (g + hf^2 + if^3 + jf^4) / (1 + \delta t + \varepsilon p)$$

**Note:**  $f = \text{Inst Freq} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$

$t = \text{Temperature } (^{\circ}\text{C})$ ;  $p = \text{Pressure (dBar)}$ ;  $\delta = \text{CTcor}$ ;  $\varepsilon = \text{CPcor}$ .

REF TEMP ( $^{\circ}\text{C}$ )	BATH SAL (PSS-78)	REF COND (S/m)	INST OUTPUT (Hz)	INST COND (S/m)	COND RESIDUAL (S/m)	SAL RESIDUAL (PSS-78)
21.3862	0	0	2550.51 <sup>2</sup>	0.00014	0.00014	-
3.0326	37.6035	3.38061	5365.83	3.38070	0.00009	0.0009
5.1484	37.6053	3.58300	5488.32	3.58298	-0.00002	-0.0001
10.2509	37.6119	4.08785	5782.37	4.08757	-0.00028	-0.0028
15.2254	37.6340	4.60241	6067.22	4.60210	-0.00031	-0.0024
20.1297	37.6510	5.12736	6344.37	5.12712	-0.00024	-0.0017
25.0393	37.6673	5.66877	6617.68	5.66857	-0.00020	-0.0008
27.0118	37.6811	5.89135	6726.76	5.89124	-0.00011	0.0002

where:

**REF TEMP** = bath set-point temperature ( $^{\circ}\text{C}$ , ITS-90), measured using the reference Standard Platinum Resistance Thermometer;

**BATH SAL** = bath salinity (PSS-78), measured using the reference Laboratory Salinometer;

**REF COND** = conductivity set-point (S/m), obtained from inverted **BATH SAL**;

**INST OUTPUT** = Instrument output frequency (Hz) at **REF TEMP**;

**INST COND** = bath conductivity (S/m), measured by Instrument;

**COND RESIDUAL** = conductivity residual: **INST COND** - **REF COND**;

**SAL RESIDUAL** = salinity residual: **INST SAL** - **REF SAL**.

<sup>2</sup> In air, with a perfectly dry conductivity cell.



# Section D

## Data & Results

### CALIBRATION



## CONDUCTIVITY

Test date: 24 November 2014

Ambient conditions:

Temperature:  $21.3^{\circ}\text{C} \pm 1^{\circ}\text{C}$

Relative Humidity:  $48\% \pm 10\%$

Atmospheric pressure: 990.7 hPa

### New conductivity calibration coefficients

$$\begin{array}{ll} g = -1.0086453e+00 & \text{CPcor} = -9.570001e-08 \\ h = 1.4765360e-01 & \text{CTcor} = 3.250000e-06 \\ i = -1.5132195e-04 & \text{WBOTC} = 1.593064e-07 \\ j = 3.1584722e-05 & \end{array}$$

$$\text{Conductivity (Siemens/m)} = (g + hf^2 + if^3 + jf^4) / (1 + \delta t + \varepsilon p)$$

**Note:**  $f$  = Inst Freq \* sqrt ( $1.0 + \text{WBOTC} * t$ ) / 1000.0

$t$  = Temperature ( $^{\circ}\text{C}$ );  $p$  = Pressure (dBar);  $\delta$  = CTcor;  $\varepsilon$  = CPcor.

REF TEMP ( $^{\circ}\text{C}$ )	BATH SAL (PSS-78)	REF COND (S/m)	INST OUTPUT (Hz)	INST COND (S/m)	COND RESIDUAL (S/m)
21.1527	0	0	2615.24 <sup>3</sup>	0	0
3.0329	37.8606	3.40142	5463.01	3.40142	0
5.1473	37.6574	3.58735	5576.57	3.58735	0
10.2508	37.6605	4.09256	5873.87	4.09255	-0.00001
15.2246	37.6293	4.60181	6158.77	4.60180	-0.00001
20.1304	37.6329	5.12525	6438.23	5.12529	0.00004
25.0384	37.8890	5.69820	6730.44	5.69814	-0.00006
27.0101	37.7968	5.90716	6833.9	5.90720	0.00004

where:

**REF TEMP** = bath set-point temperature ( $^{\circ}\text{C}$ , ITS-90), measured using the reference Standard Platinum Resistance Thermometer;

**BATH SAL** = bath salinity (PSS-78), measured using the reference Laboratory Salinometer;

**REF COND** = conductivity set-point (S/m), obtained from inverted **BATH SAL**;

**INST OUTPUT** = Instrument output frequency (Hz) at **REF TEMP**;

**INST COND** = bath conductivity (S/m), measured by Instrument;

**COND RESIDUAL** = conductivity residual: **INST COND** - **REF COND**.

<sup>3</sup> In air, with a perfectly dry conductivity cell.



## CONDUCTIVITY

Test date: 25 November 2014

Ambient conditions:

Temperature:  $21.2^{\circ}\text{C} \pm 1^{\circ}\text{C}$

Relative Humidity:  $43\% \pm 10\%$

Atmospheric pressure: 989.6 hPa

### New conductivity calibration coefficients

$$\begin{array}{ll} g = -9.8102801\text{e-01} & \text{CPcor} = -9.570001\text{e-08} \\ h = 1.5039891\text{e-01} & \text{CTcor} = 3.250000\text{e-06} \\ i = 1.2384635\text{e-04} & \text{WBOTC} = -2.352400\text{e-07} \\ j = 1.4582308\text{e-05} & \end{array}$$

$$\text{Conductivity (Siemens/m)} = (g + hf^2 + if^3 + jf^4) / (1 + \delta t + \varepsilon p)$$

**Note:**  $f = \text{Inst Freq} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$

$t = \text{Temperature } (^{\circ}\text{C})$ ;  $p = \text{Pressure (dBar)}$ ;  $\delta = \text{CTcor}$ ;  $\varepsilon = \text{CPcor}$ .

REF TEMP ( $^{\circ}\text{C}$ )	BATH SAL (PSS-78)	REF COND (S/m)	INST OUTPUT (Hz)	INST COND (S/m)	COND RESIDUAL (S/m)
21.3862	0	0	2550.51 <sup>4</sup>	0	0
3.0329	37.7677	3.39391	5374.06	3.39392	0.00001
5.1474	37.7622	3.59632	5496.36	3.59633	0.00001
10.2510	37.7567	4.10192	5790.36	4.10187	-0.00005
15.2245	37.8602	4.62694	6080.42	4.62696	0.00002
20.1304	37.8676	5.15361	6357.81	5.15365	0.00004
25.0380	37.7626	5.68132	6623.81	5.68126	-0.00006
27.0096	37.7659	5.90283	6732.32	5.90286	0.00003

where:

**REF TEMP** = bath set-point temperature ( $^{\circ}\text{C}$ , ITS-90), measured using the reference Standard Platinum Resistance Thermometer;

**BATH SAL** = bath salinity (PSS-78), measured using the reference Laboratory Salinometer;

**REF COND** = conductivity set-point (S/m), obtained from inverted **BATH SAL**;

**INST OUTPUT** = Instrument output frequency (Hz) at **REF TEMP**;

**INST COND** = bath conductivity (S/m), measured by Instrument;

**COND RESIDUAL** = conductivity residual: **INST COND** - **REF COND**.

<sup>4</sup> In air, with a perfectly dry conductivity cell.



## Section E

### Summary of work done

Two SBE 41 ALACE-CP CTD modules, identified by the serial numbers 2754 and 4522, were detached from their respective ARVOR floats. The T and C sensors of the modules were rinsed perfunctorily with deionized water and subjected to a post-deployment evaluation of their performances.


**The results of the evaluation showed that all the sensors were functioning properly, in line with the float manufacturer's specifications for them.**

Following the evaluation, the conductivity sensors of both the units were cleaned intensively following recommended practice, and therefore re-calibrated to account for any possible changes ensuing from this operation.

The calibration settings of the conductivity sensors of the two units were updated, and the floats were reconstituted to return them to their original state.

---

Measurements performed by: Nevio Medeot, Rajesh Nair.

Approved by:   
N. Medeot, CTO Unit.

The reported results are to be considered valid only for the specified instrument/s or sensor/s and the declared test conditions. This document is confidential; access to this document, or parts thereof, in any form is restricted to authorised persons only.