



Sea-Bird Scientific
 13431 NE 20th Street
 Bellevue, WA 98005
 USA

+1 425-643-9866
 seabird@seabird.com
 www.seabird.com

SENSOR SERIAL NUMBER: 9769
 CALIBRATION DATE: 11-Sep-17

SBE 41 CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.015591e+000 CPcor = -9.5700e-008
 h = 1.378647e-001 CTcor = 3.2500e-006
 i = -3.939431e-004 WBOTC = 6.7066e-006
 j = 4.719591e-005

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2721.08	0.00000	0.00000
1.0000	34.8032	2.97495	5394.81	2.97496	0.00000
4.5000	34.7835	3.28193	5597.90	3.28193	-0.00000
15.0000	34.7411	4.26336	6201.91	4.26336	-0.00000
18.5000	34.7319	4.60838	6400.39	4.60837	-0.00001
23.9940	34.7222	5.16555	6708.20	5.16556	0.00001
29.0000	34.7171	5.68790	6984.06	5.68791	0.00001
32.5000	34.7142	6.06020	7173.96	6.06019	-0.00001

$f = \text{Instrument Output(Hz)} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$

t = temperature (°C); p = pressure (decibars); $\delta = \text{CTcor}$; $\epsilon = \text{CPcor}$;

$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

