



SEA-BIRD
SCIENTIFIC

SBE Sea-Bird
Electronics

Sea-Bird Electronics
13431 NE 20th Street
Bellevue, Washington
98005 USA

Tel: +1 425-643-9866
seabird@seabird.com
www.seabird.com

SBE41-CP ALACE

Instrument Configuration

Instrument Serial Number: 41-8078
Instrument Firmware Version: V 7.2.5
Zero Conductivity Frequency: 2571.77
Communications Format: RS232
Communications Settings: 9600 baud, 8 Data Bits, No Parity

Installed Devices/Sensors

<i>Data Format</i>	<i>Measurement</i>	<i>Sensor Type</i>	<i>Serial Number</i>	<i>Rating</i>
Count	Temperature	Internal	N/A	N/A
Frequency	Conductivity	Internal	N/A	N/A
Count	Pressure	Kistler	4713534	2000m(2000 dBar)

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 8078
CALIBRATION DATE: 17-Jan-16

SBE 41 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:


a0 = -7.181116e-004
a1 = 2.736868e-004
a2 = -2.606311e-006
a3 = 1.255255e-007

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	14430681.2	1.0001	0.0001
4.5000	12311649.6	4.4999	-0.0001
15.0000	7797719.3	15.0001	0.0001
18.5000	6738681.5	18.5000	0.0000
23.9940	5390921.8	23.9939	-0.0001
29.0000	4426033.5	29.0000	0.0000
32.5001	3868850.0	32.5001	0.0000

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature



Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 8078
CALIBRATION DATE: 17-Jan-16

SBE 41 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.860342e-001
h = 1.496829e-001
i = -3.582180e-004
j = 4.877399e-005

CPcor = -9.5700e-008
CTcor = 3.2500e-006
WBOTC = -3.5640e-007


BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2571.77	0.00000	0.00000
1.0000	34.7687	2.97228	5151.94	2.97229	0.00000
4.5000	34.7494	3.27903	5347.32	3.27903	-0.00000
15.0000	34.7072	4.25964	5928.15	4.25964	-0.00000
18.5000	34.6982	4.60439	6118.99	4.60439	-0.00000
23.9940	34.6884	5.16108	6414.87	5.16108	0.00000
29.0000	34.6834	5.68300	6680.04	5.68300	0.00000
32.5001	34.6810	6.05507	6862.63	6.05507	-0.00000

$f = \text{Instrument Output(Hz)} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$

t = temperature (°C); p = pressure (decibars); $\delta = \text{CTcor}$; $\epsilon = \text{CPcor}$;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity



Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 8078
 CALIBRATION DATE: 15-Jan-16

SBE 41 PRESSURE CALIBRATION DATA
 2900 psia S/N 4713534

COEFFICIENTS:

PA0 = -1.632532e-001	PTCA0 = 4.909268e+004
PA1 = 3.871478e-004	PTCA1 = -2.006342e+002
PA2 = 9.635964e-014	PTCA2 = 8.182618e+000
PTHA0 = 3.273775e+002	PTCB0 = 1.055439e+002
PTHA1 = -9.480902e-005	PTCB1 = -6.736971e-003
PTHA2 = 3.349815e-012	PTCB2 = 0.000000e+000

PRESSURE SPAN CALIBRATION

THERMAL CORRECTION

PRESSURE (PSIA)	INSTRUMENT OUTPUT (counts)	THERMISTOR OUTPUT (counts)	COMPUTED PRESSURE (PSIA)	RESIDUAL (%FSR)	TEMP (°C)	THERMISTOR OUTPUT (counts)	INSTRUMENT OUTPUT (counts)
14.61	87179.6	3703456.8	14.77	0.01	32.50	3557364.60	90774.60
592.11	1575740.4	3703057.0	592.10	-0.00	29.00	3606622.20	89748.39
1169.51	3063624.1	3702712.4	1169.60	0.00	23.99	3678055.40	88559.99
1746.86	4550235.1	3702423.0	1747.04	0.01	18.50	3756465.60	87756.37
2324.23	6035713.8	3702117.2	2324.47	0.01	15.00	3806885.20	87492.51
2901.55	7519004.4	3702013.4	2901.47	-0.00	4.50	3959174.20	87921.92
2324.17	6034572.7	3702208.6	2324.02	-0.01	1.00	4011080.00	88479.70
1746.81	4549366.3	3702293.6	1746.70	-0.00			
1169.31	3062687.3	3702398.2	1169.24	-0.00			
591.76	1574278.1	3702519.8	591.53	-0.01			
14.61	86702.8	3702640.2	14.58	-0.00			
					TEMPERATURE (°C)	SPAN (mV)	
					-4.84	105.58	
					35.93	105.30	

y = thermistor output (counts)

$$t = PTHA0 + PTHA1 * y + PTHA2 * y^2$$

$$x = \text{instrument output} - PTCA0 - PTCA1 * t - PTCA2 * t^2$$

$$n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * t^2)$$

$$\text{pressure (PSIA)} = PA0 + PA1 * n + PA2 * n^2$$

$$\text{Residual (\%FSR)} = (\text{computed pressure} - \text{true pressure}) * 100 / \text{Full Scale Range}$$

Date, Offset (%FSR)

● 15-Jan-16 -0.00

