



SEA-BIRD
SCIENTIFIC

SBE Sea-Bird
Electronics

Sea-Bird Electronics
13431 NE 20th Street
Bellevue, Washington
98005 USA

Tel: +1 425-643-9866
seabird@seabird.com
www.seabird.com

SBE41-CP ALACE

Instrument Configuration

Instrument Serial Number: 41-8687
Instrument Firmware Version: V 7.2.5
Zero Conductivity Frequency: 2720.16
Communications Format: RS232
Communications Settings: 9600 baud, 8 Data Bits, No Parity

Installed Devices/Sensors

<i>Data Format</i>	<i>Measurement</i>	<i>Sensor Type</i>	<i>Serial Number</i>	<i>Rating</i>
Count	Temperature	Internal	N/A	N/A
Frequency	Conductivity	Internal	N/A	N/A
Count	Pressure	Druck	10173199	2000m(2000 dBar)

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 8687
CALIBRATION DATE: 31-Jul-16

SBE 41 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:


a0 = -8.609601e-004
a1 = 2.973702e-004
a2 = -4.084040e-006
a3 = 1.569618e-007

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	15099561.7	1.0000	-0.0000
4.5000	12888642.8	4.5000	0.0000
15.0000	8174554.3	15.0000	-0.0000
18.5000	7067378.2	18.5000	-0.0000
23.9940	5657564.5	23.9941	0.0001
29.0000	4647734.8	28.9999	-0.0001
32.5000	4064285.1	32.5000	0.0000

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature



Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 8687
 CALIBRATION DATE: 31-Jul-16

SBE 41 CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.006129e+000
 h = 1.365544e-001
 i = -3.239519e-004
 j = 4.151111e-005

CPcor = -9.5700e-008
 CTcor = 3.2500e-006
 WBOTC = -1.3478e-006


BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2720.16	0.00000	0.00000
1.0000	34.8159	2.97593	5410.76	2.97593	-0.00001
4.5000	34.7964	3.28303	5615.01	3.28304	0.00001
15.0000	34.7547	4.26486	6222.38	4.26486	0.00000
18.5000	34.7460	4.61005	6421.99	4.61003	-0.00002
23.9940	34.7366	5.16746	6731.54	5.16747	0.00001
29.0000	34.7315	5.68999	7008.95	5.68999	0.00000
32.5000	34.7279	6.06232	7199.88	6.06231	-0.00000

$f = \text{Instrument Output(Hz)} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$

t = temperature (°C); p = pressure (decibars); $\delta = \text{CTcor}$; $\epsilon = \text{CPcor}$;

$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$

$\text{Residual (Siemens/meter)} = \text{instrument conductivity} - \text{bath conductivity}$



Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 8687
 CALIBRATION DATE: 29-Jul-16

SBE 41 PRESSURE CALIBRATION DATA
 2900 psia S/N 10173199

COEFFICIENTS:

PA0 =	4.065309e-001	PTCA0 =	6.902344e+003
PA1 =	3.929651e-004	PTCA1 =	5.356709e+001
PA2 =	-3.138668e-013	PTCA2 =	-3.518171e-001
PTHA0 =	3.055490e+002	PTCB0 =	2.516563e+001
PTHA1 =	-6.025287e-005	PTCB1 =	3.250000e-004
PTHA2 =	-1.358444e-012	PTCB2 =	0.000000e+000

PRESSURE SPAN CALIBRATION

THERMAL CORRECTION

PRESSURE (PSIA)	INSTRUMENT OUTPUT (counts)	THERMISTOR OUTPUT (counts)	COMPUTED PRESSURE (PSIA)	RESIDUAL (%FSR)	TEMP (°C)	THERMISTOR OUTPUT (counts)	INSTRUMENT OUTPUT (counts)
14.61	44197.8	4281086.2	14.65	0.00	32.50	4144458.40	46291.70
591.04	1513481.7	4279551.8	591.15	0.00	29.00	4193370.40	46244.94
1167.38	2985710.1	4278493.8	1167.44	0.00	23.99	4263120.00	46090.62
1743.61	4461414.6	4277546.0	1743.72	0.00	18.50	4339506.00	45823.40
2319.85	5940343.5	4276476.8	2319.90	0.00	15.00	4388050.60	45627.93
2896.11	7422751.0	4275552.6	2896.05	-0.00	4.50	4533136.80	45218.80
2319.84	5940260.9	4275723.4	2319.86	0.00	1.00	4581308.60	45005.60
1743.59	4461116.3	4275841.6	1743.60	0.00			
1167.35	2985098.2	4275919.2	1167.19	-0.01			
590.91	1512362.0	4276092.8	590.70	-0.01			
14.60	44179.4	4275719.0	14.64	0.00			
					TEMPERATURE (°C)	SPAN (mV)	
					-5.00	25.16	
					35.00	25.18	

y = thermistor output (counts)

$$t = PTHA0 + PTHA1 * y + PTHA2 * y^2$$

$$x = \text{instrument output} - PTCA0 - PTCA1 * t - PTCA2 * t^2$$

$$n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * t^2)$$

$$\text{pressure (PSIA)} = PA0 + PA1 * n + PA2 * n^2$$

$$\text{Residual (\%FSR)} = (\text{computed pressure} - \text{true pressure}) * 100 / \text{Full Scale Range}$$

Date, Offset (%FSR)

● 29-Jul-16 0.00

