



SEA-BIRD
SCIENTIFIC

SBE Sea-Bird
Electronics

Sea-Bird Electronics
13431 NE 20th Street
Bellevue, Washington
98005 USA

Tel: +1 425-643-9866
seabird@seabird.com
www.seabird.com

SBE41-CP ALACE

Instrument Configuration

Instrument Serial Number: 41-9092
Instrument Firmware Version: V 7.2.5
Zero Conductivity Frequency: 2577.79
Communications Format: RS232
Communications Settings: 9600 baud, 8 Data Bits, No Parity

Installed Devices/Sensors

<i>Data Format</i>	<i>Measurement</i>	<i>Sensor Type</i>	<i>Serial Number</i>	<i>Rating</i>
Count	Temperature	Internal	N/A	N/A
Frequency	Conductivity	Internal	N/A	N/A
Count	Pressure	Kistler	4940442	2000m(2000 dBar)

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9092
CALIBRATION DATE: 21-Oct-16

SBE 41 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:


a0 = -8.071831e-004
a1 = 2.877650e-004
a2 = -3.520043e-006
a3 = 1.432282e-007

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	15789986.8	1.0000	0.0000
4.5000	13464087.0	4.5000	-0.0000
15.0000	8514492.1	15.0000	0.0000
18.5000	7354462.8	18.5000	-0.0000
23.9940	5879110.9	23.9940	0.0000
29.0000	4823738.0	28.9999	-0.0001
32.5001	4214626.3	32.5001	0.0000

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature



Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9092
 CALIBRATION DATE: 21-Oct-16

SBE 41 CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.000889e+000
 h = 1.512352e-001
 i = -3.630236e-004
 j = 4.931893e-005

CPcor = -9.5700e-008
 CTcor = 3.2500e-006
 WBOTC = -1.2606e-006


BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2577.79	0.00000	0.00000
1.0000	34.8617	2.97948	5139.80	2.97948	0.00000
4.5000	34.8421	3.28692	5334.11	3.28692	-0.00000
15.0000	34.8001	4.26984	5911.92	4.26983	-0.00000
18.5000	34.7914	4.61542	6101.82	4.61542	-0.00001
23.9940	34.7819	5.17345	6396.28	5.17347	0.00001
29.0000	34.7770	5.69661	6660.18	5.69661	-0.00000
32.5001	34.7749	6.06960	6841.92	6.06959	-0.00000

$f = \text{Instrument Output(Hz)} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$

t = temperature (°C); p = pressure (decibars); $\delta = \text{CTcor}$; $\epsilon = \text{CPcor}$;

$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$

$\text{Residual (Siemens/meter)} = \text{instrument conductivity} - \text{bath conductivity}$



Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 9092
 CALIBRATION DATE: 18-Oct-16

SBE 41 PRESSURE CALIBRATION DATA
 2900 psia S/N 4940442

COEFFICIENTS:

PA0 = -6.247949e-001	PTCA0 = -6.429550e+003
PA1 = 3.916486e-004	PTCA1 = -2.533008e+002
PA2 = 1.147450e-013	PTCA2 = 8.301950e+000
PTHA0 = 3.187078e+002	PTCB0 = 1.036676e+002
PTHA1 = -9.005338e-005	PTCB1 = -8.929476e-003
PTHA2 = 2.733559e-012	PTCB2 = 0.000000e+000

PRESSURE SPAN CALIBRATION

THERMAL CORRECTION

PRESSURE (PSIA)	INSTRUMENT OUTPUT (counts)	THERMISTOR OUTPUT (counts)	COMPUTED PRESSURE (PSIA)	RESIDUAL (%FSR)	TEMP (°C)	THERMISTOR OUTPUT (counts)	INSTRUMENT OUTPUT (counts)
14.61	31015.4	3718316.4	14.69	0.00	32.50	3563820.80	35282.60
590.43	1497812.6	3715719.2	590.50	0.00	29.00	3613186.40	34397.98
1166.35	2963529.2	3714708.4	1166.39	0.00	23.99	3684894.20	33474.05
1742.31	4428401.2	3713823.8	1742.44	0.00	18.50	3763680.00	32909.00
2318.24	5891686.0	3713011.8	2318.37	0.00	15.00	3814113.60	32816.18
2894.16	7353116.7	3712076.2	2894.07	-0.00	4.50	3966726.60	33774.52
2318.29	5891450.0	3711866.2	2318.29	0.00	1.00	4018093.80	34520.99
1742.20	4427586.0	3711513.8	1742.14	-0.00			
1166.46	2963204.7	3711204.8	1166.27	-0.01			
590.21	1496858.5	3710965.6	590.13	-0.00	TEMPERATURE (°C)	SPAN (mV)	
14.61	30783.8	3710210.6	14.58	-0.00	-3.82	103.70	
					35.66	103.35	

y = thermistor output (counts)

$$t = PTHA0 + PTHA1 * y + PTHA2 * y^2$$

$$x = \text{instrument output} - PTCA0 - PTCA1 * t - PTCA2 * t^2$$

$$n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * t^2)$$

$$\text{pressure (PSIA)} = PA0 + PA1 * n + PA2 * n^2$$

$$\text{Residual (\%FSR)} = (\text{computed pressure} - \text{true pressure}) * 100 / \text{Full Scale Range}$$

Date, Offset (%FSR)

● 18-Oct-16 0.00

