

Certificate no: 4330_2975_00148673
Foil batch no: 1517M

Product: 4330
Calibration date: 03.06.2018

Serial no: 2975
Page 1 of 2

Index	Temperature reference(°C)	[O2] Reference(µM)	Temperature raw data(mV)	Phase reading(°)
0	30.303	1.84	-55.493	59.83
1	20.145	1.61	270.480	60.76
2	10.255	1.61	585.200	61.53
3	0.876	1.96	858.500	62.18
4	1.021	19.90	854.540	59.70
5	1.139	42.38	851.333	56.94
6	1.256	65.51	848.087	54.41
7	1.336	107.31	845.947	50.47
8	1.381	150.05	844.687	47.12
9	1.402	213.60	844.113	43.07
10	1.449	316.30	842.847	38.21
11	1.463	419.11	842.487	34.64
12	1.467	530.77	842.373	31.74
13	10.783	18.55	568.873	58.23
14	10.602	35.02	574.487	55.40
15	10.513	52.00	577.220	52.86
16	10.470	84.94	578.553	48.67
17	10.423	117.50	580.000	45.31
18	10.399	170.45	580.733	41.00
19	10.387	253.73	581.120	36.06
20	10.392	337.02	580.953	32.55
21	10.398	421.83	580.780	29.89
22	20.409	15.44	261.947	57.03
23	20.273	28.21	266.367	54.04
24	20.198	42.00	268.787	51.27
25	20.155	68.49	270.160	46.83
26	20.136	94.72	270.793	43.31
27	20.130	136.16	271.000	39.00
28	20.140	204.17	270.667	34.03
29	20.154	271.99	270.200	30.59
30	20.159	341.29	270.033	28.02
31	30.149	13.59	-50.673	55.73
32	30.146	23.60	-50.607	52.64
33	30.165	34.54	-51.193	49.73
34	30.192	56.57	-51.980	45.01
35	30.231	78.29	-53.220	41.35
36	30.247	112.68	-53.727	37.00
37	30.274	168.55	-54.540	32.13
38	30.299	222.82	-55.353	28.89
39	30.311	283.45	-55.700	26.34

Certificate no: 4330_2975_00148673
Foil batch no: 1517M

Product: 4330
Calibration date: 03.06.2018

Serial no: 2975
Page 2 of 2

Index	0	1	2	3	4	5	6
SVUFoilCoef	2.73262E-03	1.19088E-04	2.21040E-06	2.00474E02	-2.44286E-01	-4.13361E01	3.95887E00
TempCoef	2.85377E01	-3.16139E-02	3.49146E-06	-4.88527E-09	0.00000E00	0.00000E00	

